

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122722\
 Data File : VU052574.D
 Acq On : 27 Dec 2022 18:59
 Operator : JC/MD
 Sample : N6189-01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 COL41

Quant Time: Dec 28 01:24:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR122022WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Dec 28 01:18:16 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	185819	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	178308	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.806	152	81151	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	53900	2.863	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	57.200%
7) Chloroethane-d5	1.916	69	57982	3.571	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	71.400%
11) 1,1-Dichloroethene-d2	2.568	65	26391	3.258	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.200%
20) 2-Butanone-d5	4.646	46	318908	54.628	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	109.260%
24) Chloroform-d	5.060	84	134605	4.238	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.800%
26) 1,2-Dichloroethane-d4	5.697	65	82234	4.464	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.200%
32) Benzene-d6	5.723	84	267665	4.378	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.600%
36) 1,2-Dichloropropane-d6	6.687	67	98825	4.696	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.000%
41) Toluene-d8	7.896	98	207303	4.039	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.800%
43) trans-1,3-Dichloroprop...	8.176	79	32994	4.579	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.600%
46) 2-Hexanone-d5	8.633	63	219085	55.571	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	111.140%
56) 1,1,2,2-Tetrachloroeth...	10.751	84	85875	4.806	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.200%
66) 1,2-Dichlorobenzene-d4	12.189	152	82490	4.993	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122722\
Data File : VU052574.D
Acq On : 27 Dec 2022 18:59
Operator : JC/MD
Sample : N6189-01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
COL41

Quant Time: Dec 28 01:24:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR122022WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Dec 28 01:18:16 2022
Response via : Initial Calibration

