

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122723\
 Data File : VU057312.D
 Acq On : 28 Dec 2023 05:09
 Operator : MD/SY
 Sample : 06020-07
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BH7W8

Quant Time: Dec 28 05:27:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR122723WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Dec 28 00:43:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	244863	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	246071	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	99320	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	104018	4.941	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	98.800%
7) Chloroethane-d5	1.911	69	90115	5.330	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	106.600%
11) 1,1-Dichloroethene-d2	2.560	65	41709	4.978	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	99.600%
20) 2-Butanone-d5	4.640	46	273439	57.352	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	114.700%
24) Chloroform-d	5.055	84	188579	5.038	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.800%
26) 1,2-Dichloroethane-d4	5.695	65	102955	5.382	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.600%
32) Benzene-d6	5.721	84	350804	5.001	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.000%
36) 1,2-Dichloropropane-d6	6.682	67	117382	5.340	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	106.800%
41) Toluene-d8	7.894	98	280189	4.630	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.600%
43) trans-1,3-Dichloroprop...	8.174	79	41937	4.755	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	95.200%
46) 2-Hexanone-d5	8.631	63	200812	50.272	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.540%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	97587	5.286	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	105.800%
66) 1,2-Dichlorobenzene-d4	12.190	152	99858	5.521	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	110.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122723\
Data File : VU057312.D
Acq On : 28 Dec 2023 05:09
Operator : MD/SY
Sample : 06020-07
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BH7W8

Quant Time: Dec 28 05:27:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR122723WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Dec 28 00:43:53 2023
Response via : Initial Calibration

