

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122723\
 Data File : VU057327.D
 Acq On : 28 Dec 2023 11:35
 Operator : MD/SY
 Sample : 06020-08
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_U
ClientSampleId :
 BH7W9

Quant Time: Dec 28 20:48:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR122723WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Dec 28 05:49:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	220367	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	226497	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	89888	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	102006	5.384	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	107.600%	
7) Chloroethane-d5	1.911	69	88523	5.818	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	116.400%	
11) 1,1-Dichloroethene-d2	2.560	65	39413	5.227	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	104.600%	
20) 2-Butanone-d5	4.631	46	246607	57.474	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	114.940%	
24) Chloroform-d	5.055	84	185122	5.495	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	110.000%	
26) 1,2-Dichloroethane-d4	5.695	65	99096	5.756	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	115.200%	
32) Benzene-d6	5.721	84	333170	5.160	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	103.200%	
36) 1,2-Dichloropropane-d6	6.682	67	113295	5.600	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	112.000%	
41) Toluene-d8	7.894	98	262578	4.714	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	94.200%	
43) trans-1,3-Dichloroprop...	8.174	79	40314	4.966	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	99.400%	
46) 2-Hexanone-d5	8.631	63	186405	50.698	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	101.400%	
56) 1,1,2,2-Tetrachloroeth...	10.749	84	92465	5.442	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	108.800%	
66) 1,2-Dichlorobenzene-d4	12.190	152	95106	5.811	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	116.200%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122723\
 Data File : VU057327.D
 Acq On : 28 Dec 2023 11:35
 Operator : MD/SY
 Sample : 06020-08
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BH7W9

Quant Time: Dec 28 20:48:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR122723WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Dec 28 05:49:04 2023
 Response via : Initial Calibration

