

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122723\
 Data File : VU057302.D
 Acq On : 28 Dec 2023 01:06
 Operator : MD/SY
 Sample : 06021-10
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BH7Z2

Manual Integrations
 APPROVED

Reviewed By : John Carlone 12/29/2023
 Supervised By : Mahesh Dadoda 12/29/2023

Quant Time: Dec 28 01:50:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR122723WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Dec 28 00:43:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	255294	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	256858	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	103237	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	115825	5.277	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	105.600%	
7) Chloroethane-d5	1.911	69	98754	5.602	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	112.000%	
11) 1,1-Dichloroethene-d2	2.560	65	47447	5.432	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	108.600%	
20) 2-Butanone-d5	4.637	46	287400	57.817	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	115.640%	
24) Chloroform-d	5.055	84	206763	5.298	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	106.000%	
26) 1,2-Dichloroethane-d4	5.695	65	109852	5.508	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	110.200%	
32) Benzene-d6	5.721	84	386730	5.282	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	105.600%	
36) 1,2-Dichloropropane-d6	6.682	67	129846	5.659	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	113.200%	
41) Toluene-d8	7.891	98	314720	4.982	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	99.600%	
43) trans-1,3-Dichloroprop...	8.174	79	48365	5.254	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	105.000%	
46) 2-Hexanone-d5	8.631	63	234101	56.145	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	112.280%	
56) 1,1,2,2-Tetrachloroeth...	10.750	84	103614	5.377	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	107.600%	
66) 1,2-Dichlorobenzene-d4	12.187	152	111279	5.920	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	118.400%	
Target Compounds						Qvalue
5) Vinyl chloride	1.602	62	126095	5.917	ug/L	97
12) 1,1-Dichloroethene	2.573	96	7072	0.424	ug/L #	1
13) Acetone	2.663	43	31780m	9.809	ug/L	
14) Carbon disulfide	2.789	76	6556	0.138	ug/L	97
18) trans-1,2-Dichloroethene	3.351	96	2246	0.136	ug/L	78
19) 1,1-Dichloroethane	3.862	63	64773	1.862	ug/L	96
22) cis-1,2-Dichloroethene	4.657	96	122556	6.404	ug/L	97
29) 1,1,1-Trichloroethane	5.309	97	259762	7.985	ug/L	100
34) Trichloroethene	6.534	95	1111009	56.344	ug/L	97
42) Toluene	7.968	91	21813	0.274	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122723\
Data File : VU057302.D
Acq On : 28 Dec 2023 01:06
Operator : MD/SY
Sample : 06021-10
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 26 Sample Multiplier: 1


Instrument :
MSVOA_U
ClientSampleId :
BH7Z2

Quant Time: Dec 28 01:50:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR122723WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Dec 28 00:43:53 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 12/29/2023
Supervised By :Mahesh Dadoda 12/29/2023









#5
 Vinyl chloride
 Concen: 5.917 ug/L
 RT: 1.602 min Scan# 91
 Delta R.T. 0.000 min
 Lab File: VU057302.D
 Acq: 28 Dec 2023 01:06

Instrument :
 MSVOA_U
 ClientSampleId :
 BH7Z2

Tgt Ion: 62 Resp: 12609
 Ion Ratio Lower Upper
 62 100
 64 34.3 22.8 42.4




Manual Integrations
 APPROVED


Reviewed By :John Carlone 12/29/2023
 Supervised By :Mahesh Dadoda 12/29/2023



#12
 1,1-Dichloroethene
 Concen: 0.424 ug/L
 RT: 2.573 min Scan# 399
 Delta R.T. 0.000 min
 Lab File: VU057302.D
 Acq: 28 Dec 2023 01:06

Tgt Ion: 96 Resp: 7072
 Ion Ratio Lower Upper
 96 100
 61 305.1 122.2 226.9#
 63 1317.7 99.2 184.2#





#13

Acetone

Concen: 9.809 ug/L m

RT: 2.663 min Scan# 41

Delta R.T. 0.003 min

Lab File: VU057302.D

Acq: 28 Dec 2023 01:06

Instrument :

MSVOA_U

ClientSampleId :

BH7Z2

Tgt Ion: 43 Resp: 31780

Ion Ratio Lower Upper

43 100





58 20.0 0.0 67.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 12/29/2023

Supervised By :Mahesh Dadoda 12/29/2023



#14

Carbon disulfide

Concen: 0.138 ug/L

RT: 2.789 min Scan# 466

Delta R.T. 0.000 min

Lab File: VU057302.D




Acq: 28 Dec 2023 01:06


Tgt Ion: 76 Resp: 6556

Ion Ratio Lower Upper

76 100


78 10.2 7.3 10.9





#18
trans-1,2-Dichloroethene
Concen: 0.136 ug/L
RT: 3.351 min Scan# 640
Delta R.T. 0.003 min
Lab File: VU057302.D
Acq: 28 Dec 2023 01:06




Instrument :
MSVOA_U
ClientSampleId :
BH7Z2




Tgt Ion: 96 Resp: 2240
Ion Ratio Lower Upper
96 100
61 170.5 97.5 181.1
98 76.8 45.9 85.3

Manual Integrations
APPROVED



Reviewed By :John Carlone 12/29/2023
Supervised By :Mahesh Dadoda 12/29/2023




#19
1,1-Dichloroethane
Concen: 1.862 ug/L
RT: 3.862 min Scan# 800
Delta R.T. 0.003 min
Lab File: VU057302.D
Acq: 28 Dec 2023 01:06




Tgt Ion: 63 Resp: 64773
Ion Ratio Lower Upper
63 100
65 31.7 23.6 43.8
83 12.2 9.8 18.2





#22
 cis-1,2-Dichloroethene
 Concen: 6.404 ug/L
 RT: 4.657 min Scan# 1047
 Delta R.T. 0.000 min
 Lab File: VU057302.D
 Acq: 28 Dec 2023 01:06




Instrument :
 MSVOA_U
 ClientSampleId :
 BH7Z2




Tgt Ion: 96 Resp: 122550
 Ion Ratio Lower Upper
 96 100
 61 133.9 91.1 169.3
 98 63.1 42.6 79.0

Manual Integrations
 APPROVED



Reviewed By :John Carlone 12/29/2023
 Supervised By :Mahesh Dadoda 12/29/2023




#29
 1,1,1-Trichloroethane
 Concen: 7.985 ug/L
 RT: 5.309 min Scan# 1250
 Delta R.T. 0.000 min
 Lab File: VU057302.D
 Acq: 28 Dec 2023 01:06




Tgt Ion: 97 Resp: 259762
 Ion Ratio Lower Upper
 97 100
 99 64.2 51.7 77.5
 61 42.7 34.1 51.1





#34
Trichloroethene
Concen: 56.344 ug/L
RT: 6.534 min Scan# 1000
Delta R.T. 0.000 min
Lab File: VU057302.D
Acq: 28 Dec 2023 01:06




Instrument :
MSVOA_U
ClientSampleId :
BH7Z2



Tgt Ion: 95 Resp: 1111009
Ion Ratio Lower Upper
95 100
97 64.5 45.3 84.1
132 88.1 63.1 117.3
130 92.1 67.7 125.7

Manual Integrations
APPROVED

Reviewed By :John Carlone 12/29/2023
Supervised By :Mahesh Dadoda 12/29/2023



#42
Toluene
Concen: 0.274 ug/L
RT: 7.968 min Scan# 2077
Delta R.T. 0.006 min
Lab File: VU057302.D
Acq: 28 Dec 2023 01:06

Tgt Ion: 91 Resp: 21813
Ion Ratio Lower Upper
91 100
92 57.0 41.6 77.2

