

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122821\
 Data File : VU046621.D
 Acq On : 28 Dec 2021 18:48
 Operator : SY/MD
 Sample : M5193-12
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0J31

Quant Time: Dec 29 01:31:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR122121WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Dec 29 01:29:25 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	59340	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	63378	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	29276	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	22491	4.812	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	96.200%	
7) Chloroethane-d5	1.919	69	21188	5.757	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	115.200%	
11) 1,1-Dichloroethene-d2	2.575	65	10879	5.520	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	110.400%	
20) 2-Butanone-d5	4.646	46	82600	60.924	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	121.840%	
24) Chloroform-d	5.067	84	51880	6.058	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	121.200%	
26) 1,2-Dichloroethane-d4	5.707	65	29921	6.111	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	122.200%	
32) Benzene-d6	5.732	84	97831	5.537	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	110.800%	
36) 1,2-Dichloropropane-d6	6.694	67	30968	5.833	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	116.600%	
41) Toluene-d8	7.903	98	77785	4.851	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	97.000%	
43) trans-1,3-Dichloroprop...	8.182	79	11649	4.744	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	94.800%	
46) 2-Hexanone-d5	8.636	63	68462	47.491	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	94.980%	
56) 1,1,2,2-Tetrachloroeth...	10.758	84	29354	5.387	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	107.800%	
66) 1,2-Dichlorobenzene-d4	12.195	152	30458	5.648	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	113.000%	
Target Compounds						
					Qvalue	
3) Chloromethane	1.523	50	1313	0.168	ug/L	# 82


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122821\
Data File : VU046621.D
Acq On : 28 Dec 2021 18:48
Operator : SY/MD
Sample : M5193-12
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0J31

Quant Time: Dec 29 01:31:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR122121WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Dec 29 01:29:25 2021
Response via : Initial Calibration





#3
Chloromethane
Concen: 0.168 ug/L
RT: 1.523 min Scan# 51
Delta R.T. 0.000 min
Lab File: VU046621.D
Acq: 28 Dec 2021 18:48

Instrument :
MSVOA_U
ClientSampleId :
C0J31

Tgt Ion: 50 Resp: 1313
Ion Ratio Lower Upper
50 100
52 43.2 23.0 42.8#

