

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122823\
 Data File : VU057354.D
 Acq On : 29 Dec 2023 00:27
 Operator : MD/SY
 Sample : 05998-12
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F7A80

Quant Time: Dec 29 00:40:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR122723WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Dec 28 21:37:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	228815	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	236084	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	90166	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.599	65	97108	4.937	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	98.800%
7) Chloroethane-d5	1.911	69	87177	5.518	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	110.400%
11) 1,1-Dichloroethene-d2	2.563	65	40286	5.146	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	103.000%
20) 2-Butanone-d5	4.647	46	208247	46.742	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	93.480%
24) Chloroform-d	5.055	84	189237	5.410	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.200%
26) 1,2-Dichloroethane-d4	5.698	65	94348	5.278	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.600%
32) Benzene-d6	5.721	84	345240	5.130	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.600%
36) 1,2-Dichloropropane-d6	6.685	67	115937	5.498	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	110.000%
41) Toluene-d8	7.894	98	259594	4.471	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.400%
43) trans-1,3-Dichloroprop...	8.174	79	38284	4.525	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.400%
46) 2-Hexanone-d5	8.634	63	148013	38.622	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	77.240%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	86369	4.877	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.600%
66) 1,2-Dichlorobenzene-d4	12.187	152	88324	5.380	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	107.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122823\
Data File : VU057354.D
Acq On : 29 Dec 2023 00:27
Operator : MD/SY
Sample : 05998-12
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F7A80

Quant Time: Dec 29 00:40:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR122723WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Dec 28 21:37:03 2023
Response via : Initial Calibration

