

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U122920\
 Data File : VU041830.D
 Acq On : 29 Dec 2020 11:43
 Operator : SY/MD
 Sample : VU1229WBL01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK13

Quant Time: Dec 30 03:11:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR121220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Dec 30 01:48:43 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	159329	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	153494	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	66859	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	59753	4.80	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	96.00%
7) Chloroethane-d5	1.92	69	55193	5.72	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	114.40%
11) 1,1-Dichloroethene-d2	2.57	63	90112	3.93	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	78.60%
20) 2-Butanone-d5	4.66	46	150598	42.09	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	84.18%
24) Chloroform-d	5.08	84	103157	4.83	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.60%
26) 1,2-Dichloroethane-d4	5.72	65	57800	4.77	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.40%
32) Benzene-d6	5.74	84	205919	4.69	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.80%
36) 1,2-Dichloropropane-d6	6.70	67	63871	4.63	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.60%
41) Toluene-d8	7.90	98	178973	4.51	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.20%
43) trans-1,3-Dichloropropene-	8.19	79	26455	4.27	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.40%
46) 2-Hexanone-d5	8.64	63	143181	42.27	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	84.54%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	58249	4.65	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.00%
64) 1,2-Dichlorobenzene-d4	12.19	152	66887	5.63	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	112.60%

Target Compounds					Ovalue
16) Methylene chloride	3.06	84	3466	0.269 ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU122920\
Data File : VU041830.D
Acq On : 29 Dec 2020 11:43
Operator : SY/MD
Sample : VU1229WBL01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK13

Quant Time: Dec 30 03:11:17 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR121220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Dec 30 01:48:43 2020
Response via : Initial Calibration

