

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U122920\
 Data File : VU041833.D
 Acq On : 29 Dec 2020 13:04
 Operator : SY/MD
 Sample : L5204-04
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AM3

Quant Time: Dec 30 03:23:55 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR121220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Dec 30 01:48:43 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	151114	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	148284	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	65496	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	53981	4.57	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	91.40%
7) Chloroethane-d5	1.92	69	53436	5.84	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	116.80%
11) 1,1-Dichloroethene-d2	2.57	63	82076	3.77	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	75.40%
20) 2-Butanone-d5	4.66	46	146689	43.23	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	86.46%
24) Chloroform-d	5.08	84	97084	4.80	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.00%
26) 1,2-Dichloroethane-d4	5.71	65	56555	4.93	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.60%
32) Benzene-d6	5.74	84	194934	4.59	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.80%
36) 1,2-Dichloropropane-d6	6.70	67	60062	4.51	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	90.20%
41) Toluene-d8	7.90	98	171283	4.47	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.40%
43) trans-1,3-Dichloropropene-	8.19	79	25202	4.21	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.20%
46) 2-Hexanone-d5	8.64	63	134406	41.07	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	82.14%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	55002	4.55	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.00%
64) 1,2-Dichlorobenzene-d4	12.19	152	63232	5.43	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	108.60%

Target Compounds					Ovalue
16) Methylene chloride	3.05	84	1510	0.123 ug/L	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU122920\
Data File : VU041833.D
Acq On : 29 Dec 2020 13:04
Operator : SY/MD
Sample : L5204-04
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AM3

Quant Time: Dec 30 03:23:55 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR121220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Dec 30 01:48:43 2020
Response via : Initial Calibration

