

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122921\
 Data File : VU046628.D
 Acq On : 29 Dec 2021 10:37
 Operator : SY/MD
 Sample : VU1229WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK049

Quant Time: Dec 30 04:22:08 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR122121WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Dec 30 04:21:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)

Internal Standards					
1) 1,4-Difluorobenzene	6.253	114	65138	5.000 ug/L	0.00
28) Chlorobenzene-d5	9.420	117	70566	5.000 ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.816	152	33029	5.000 ug/L	0.00

System Monitoring Compounds					
4) Vinyl Chloride-d3	1.604	65	32291	6.294 ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 125.800%	
7) Chloroethane-d5	1.919	69	22972	5.686 ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 113.800%	
11) 1,1-Dichloroethene-d2	2.575	65	13126	6.067 ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 121.400%	
20) 2-Butanone-d5	4.652	46	79534	53.441 ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	= 106.880%	
24) Chloroform-d	5.070	84	50802	5.404 ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 108.000%	
26) 1,2-Dichloroethane-d4	5.706	65	29920	5.567 ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 111.400%	
32) Benzene-d6	5.732	84	96731	4.917 ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 98.400%	
36) 1,2-Dichloropropane-d6	6.694	67	30477	5.156 ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 103.200%	
41) Toluene-d8	7.903	98	80276	4.496 ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 90.000%	
43) trans-1,3-Dichloroprop...	8.182	79	13104	4.793 ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 95.800%	
46) 2-Hexanone-d5	8.639	63	65182	40.610 ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 81.220%	
56) 1,1,2,2-Tetrachloroeth...	10.761	84	28277	4.661 ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 93.200%	
66) 1,2-Dichlorobenzene-d4	12.195	152	29461	4.842 ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 96.800%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122921\
Data File : VU046628.D
Acq On : 29 Dec 2021 10:37
Operator : SY/MD
Sample : VU1229WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK049

Quant Time: Dec 30 04:22:08 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR122121WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Dec 30 04:21:35 2021
Response via : Initial Calibration

