

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122921\
 Data File : VU046630.D
 Acq On : 29 Dec 2021 15:57
 Operator : SY/MD
 Sample : M5170-12
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Dec 30 04:22:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR122121WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Dec 30 04:21:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	61527	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	66415	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.815	152	30204	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	31447	6.489	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 129.800%		
7) Chloroethane-d5	1.919	69	22618	5.927	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 118.600%		
11) 1,1-Dichloroethene-d2	2.575	65	12586	6.159	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 123.200%		
20) 2-Butanone-d5	4.649	46	70215	49.949	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	= 99.900%		
24) Chloroform-d	5.070	84	50106	5.642	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 112.800%		
26) 1,2-Dichloroethane-d4	5.706	65	28588	5.631	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 112.600%		
32) Benzene-d6	5.732	84	93269	5.038	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 100.800%		
36) 1,2-Dichloropropane-d6	6.693	67	29390	5.283	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 105.600%		
41) Toluene-d8	7.902	98	74401	4.428	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 88.600%		
43) trans-1,3-Dichloroprop...	8.182	79	11868	4.613	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 92.200%		
46) 2-Hexanone-d5	8.639	63	58483	38.714	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 77.420%		
56) 1,1,2,2-Tetrachloroeth...	10.758	84	27266	4.775	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 95.600%		
66) 1,2-Dichlorobenzene-d4	12.195	152	27296	4.906	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 98.200%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU122921\
Data File : VU046630.D
Acq On : 29 Dec 2021 15:57
Operator : SY/MD
Sample : M5170-12
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Dec 30 04:22:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR122121WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Dec 30 04:21:35 2021
Response via : Initial Calibration

