

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U031220\
 Data File : VU037211.D
 Acq On : 13 Mar 2020 00:13
 Operator : JC/MD
 Sample : L1886-02
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0NA2

Quant Time: Mar 13 08:49:57 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM031220WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Mar 13 08:28:33 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	390548	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	384946	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	161602	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	139235	46.06	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	92.12%
7) Chloroethane-d5	1.93	69	117720	48.31	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	96.62%
11) 1,1-Dichloroethene-d2	2.59	63	205455	37.05	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	74.10%
21) 2-Butanone-d5	4.67	46	255678	106.89	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	106.89%
24) Chloroform-d	5.11	84	283690	48.82	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.64%
26) 1,2-Dichloroethane-d4	5.74	65	211805	49.79	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.58%
32) Benzene-d6	5.76	84	519129	48.99	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.98%
36) 1,2-Dichloropropane-d6	6.72	67	173005	49.87	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.74%
41) Toluene-d8	7.92	98	481279	47.51	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.02%
43) trans-1,3-Dichloropropene-	8.20	79	87538	46.36	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.72%
47) 2-Hexanone-d5	8.66	63	179473	99.93	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	99.93%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	254038	47.28	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	94.56%
64) 1,2-Dichlorobenzene-d4	12.21	152	179919	53.67	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	107.34%

Target Compounds

				Ovalue	
12) 1,1-Dichloroethene	2.60	96	12097	5.249 ug/L	# 1
13) Acetone	2.66	43	5246	2.312 ug/L	99
20) cis-1,2-Dichloroethene	4.71	96	17440	5.971 ug/L	97
30) 1,1,1-Trichloroethane	5.36	97	17958	3.515 ug/L	98
34) Trichloroethene	6.57	95	223049	75.651 ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU031220\
Data File : VU037211.D
Acq On : 13 Mar 2020 00:13
Operator : JC/MD
Sample : L1886-02
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0NA2

Quant Time: Mar 13 08:49:57 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM031220WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Mar 13 08:28:33 2020
Response via : Initial Calibration

