

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU031220\
 Data File : VU037193.D
 Acq On : 12 Mar 2020 16:59
 Operator : JC/MD
 Sample : VIBLK50
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK50

Quant Time: Mar 13 07:50:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM031220WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Mar 13 07:30:58 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	410854	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	396209	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	172807	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	149378	46.98	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	93.96%
7) Chloroethane-d5	1.93	69	123374	48.12	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	96.24%
11) 1,1-Dichloroethene-d2	2.59	63	214675	36.80	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.60%
21) 2-Butanone-d5	4.67	46	263838	104.85	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	104.85%
24) Chloroform-d	5.11	84	288557	47.20	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.40%
26) 1,2-Dichloroethane-d4	5.74	65	222289	49.67	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.34%
32) Benzene-d6	5.76	84	538027	49.33	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.66%
36) 1,2-Dichloropropane-d6	6.72	67	180957	50.68	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.36%
41) Toluene-d8	7.92	98	499963	47.95	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.90%
43) trans-1,3-Dichloropropene-	8.20	79	91674	47.17	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	94.34%
47) 2-Hexanone-d5	8.66	63	183057	99.03	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	99.03%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	260510	47.11	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	94.22%
64) 1,2-Dichlorobenzene-d4	12.21	152	186677	52.08	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	104.16%


Target Compounds					Ovalue
34) Trichloroethene	6.57	95	4285	1.412 ug/L	84


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU031220\
Data File : VU037193.D
Acq On : 12 Mar 2020 16:59
Operator : JC/MD
Sample : VIBLK50
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK50

Quant Time: Mar 13 07:50:56 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM031220WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Mar 13 07:30:58 2020
Response via : Initial Calibration





#34

Trichloroethene

Concen: 1.412 ug/L

RT: 6.57 min Scan# 1628

Delta R.T. 0.00 min

Lab File: VU037193.D

Acq: 12 Mar 2020 16:59

Instrument :

MSVOA_U

ClientSampleId :

VIBLK50

Tot Ion: 95 Resp: 4285



Ion Ratio Lower Upper

95 100

97 51.3 43.8 81.4

132 74.7 62.6 116.3

130 108.4 64.8 120.4



Abundance Ion 95.00 (94.70 to 95.70): VU0

Ion 97.00 (96.70 to 97.70): VU0

Ion 132.00 (131.70 to 132.70): V

Ion 130.00 (129.70 to 130.70): V

