

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU031220\
 Data File : VU037200.D
 Acq On : 12 Mar 2020 19:39
 Operator : JC/MD
 Sample : L1901-03
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0NC5

Quant Time: Mar 13 08:20:43 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM031220WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Mar 13 07:30:58 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	397800	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	392081	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	165828	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	140843	45.75	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	91.50%
7) Chloroethane-d5	1.93	69	118597	47.78	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	95.56%
11) 1,1-Dichloroethene-d2	2.59	63	200682	35.53	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.06%
21) 2-Butanone-d5	4.67	46	256323	105.20	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	105.20%
24) Chloroform-d	5.11	84	280516	47.39	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.78%
26) 1,2-Dichloroethane-d4	5.74	65	210715	48.63	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.26%
32) Benzene-d6	5.76	84	516409	47.85	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.70%
36) 1,2-Dichloropropane-d6	6.72	67	170494	48.26	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.52%
41) Toluene-d8	7.92	98	475297	46.07	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.14%
43) trans-1,3-Dichloropropene-	8.20	79	86712	45.09	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.18%
47) 2-Hexanone-d5	8.66	63	174819	95.57	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	95.57%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	248649	45.44	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	90.88%
64) 1,2-Dichlorobenzene-d4	12.21	152	177647	51.65	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.30%


Target Compounds					Ovalue
13) Acetone	2.65	43	13333	5.768 ug/L	92


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU031220\
Data File : VU037200.D
Acq On : 12 Mar 2020 19:39
Operator : JC/MD
Sample : L1901-03
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0NC5

Quant Time: Mar 13 08:20:43 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM031220WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Mar 13 07:30:58 2020
Response via : Initial Calibration





#13
 Acetone
 Concen: 5.768 ug/L
 RT: 2.65 min Scan# 408
 Delta R.T. 0.00 min
 Lab File: VU037200.D
 Acq: 12 Mar 2020 19:39

Instrument :
 MSVOA_U
 ClientSampleId :
 CONC5

Tot Ion: 43 Resp: 13333
 Ion Ratio Lower Upper
 43 100
 58 24.5 0.0 57.2

