

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU031220\
 Data File : VU037205.D
 Acq On : 12 Mar 2020 21:56
 Operator : JC/MD
 Sample : VU0312WBL02
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK43

Quant Time: Mar 13 08:35:31 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM031220WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Mar 13 08:28:33 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	390761	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	383192	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	159674	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	146090	48.31	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	96.62%
7) Chloroethane-d5	1.93	69	124140	50.91	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	101.82%
11) 1,1-Dichloroethene-d2	2.59	63	207793	37.45	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	74.90%
21) 2-Butanone-d5	4.67	46	265279	110.84	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	110.84%
24) Chloroform-d	5.11	84	281541	48.42	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.84%
26) 1,2-Dichloroethane-d4	5.74	65	220940	51.91	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.82%
32) Benzene-d6	5.76	84	534227	50.65	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.30%
36) 1,2-Dichloropropane-d6	6.72	67	175680	50.88	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.76%
41) Toluene-d8	7.92	98	493080	48.90	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.80%
43) trans-1,3-Dichloropropene-	8.20	79	89744	47.75	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	95.50%
47) 2-Hexanone-d5	8.66	63	182636	102.16	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	102.16%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	256268	47.91	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.82%
64) 1,2-Dichlorobenzene-d4	12.21	152	178548	53.91	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	107.82%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU031220\
Data File : VU037205.D
Acq On : 12 Mar 2020 21:56
Operator : JC/MD
Sample : VU0312WBL02
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK43

Quant Time: Mar 13 08:35:31 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM031220WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Mar 13 08:28:33 2020
Response via : Initial Calibration

