

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042322\
 Data File : VU048248.D
 Acq On : 23 Apr 2022 16:02
 Operator : SY/MD
 Sample : N2488-07
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0HY9

Quant Time: Apr 25 01:26:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Apr 25 01:24:32 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	295148	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	303756	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	140487	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	107287	50.127	ug/L	0.00
Spiked Amount 50.000	Range 60	- 135	Recovery	=	100.260%	
7) Chloroethane-d5	1.890	69	85437	50.625	ug/L	0.00
Spiked Amount 50.000	Range 70	- 130	Recovery	=	101.260%	
11) 1,1-Dichloroethene-d2	2.562	63	134417	34.326	ug/L	0.00
Spiked Amount 50.000	Range 60	- 125	Recovery	=	68.660%	
21) 2-Butanone-d5	4.623	46	179427	97.670	ug/L	0.00
Spiked Amount 100.000	Range 40	- 130	Recovery	=	97.670%	
24) Chloroform-d	5.063	84	188340	48.093	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	96.180%	
26) 1,2-Dichloroethane-d4	5.703	65	119752	49.858	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	99.720%	
32) Benzene-d6	5.726	84	398813	52.178	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	104.360%	
36) 1,2-Dichloropropane-d6	6.690	67	131736	56.121	ug/L	0.00
Spiked Amount 50.000	Range 70	- 120	Recovery	=	112.240%	
41) Toluene-d8	7.899	98	319654	42.925	ug/L	0.00
Spiked Amount 50.000	Range 80	- 120	Recovery	=	85.840%	
43) trans-1,3-Dichloroprop...	8.179	79	46601	39.187	ug/L	0.00
Spiked Amount 50.000	Range 60	- 125	Recovery	=	78.380%	
47) 2-Hexanone-d5	8.639	63	125816	91.560	ug/L	0.00
Spiked Amount 100.000	Range 45	- 130	Recovery	=	91.560%	
56) 1,1,2,2-Tetrachloroeth...	10.758	84	214230	53.179	ug/L	0.00
Spiked Amount 50.000	Range 65	- 120	Recovery	=	106.360%	
66) 1,2-Dichlorobenzene-d4	12.195	152	130995	52.411	ug/L	0.00
Spiked Amount 50.000	Range 80	- 120	Recovery	=	104.820%	
Target Compounds						Qvalue
51) Chlorobenzene	9.449	112	3851	0.588	ug/L	91
65) 1,4-Dichlorobenzene	11.838	146	3376	0.772	ug/L	97
67) 1,2-Dichlorobenzene	12.214	146	4162	0.986	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042322\
Data File : VU048248.D
Acq On : 23 Apr 2022 16:02
Operator : SY/MD
Sample : N2488-07
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0HY9

Quant Time: Apr 25 01:26:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Apr 25 01:24:32 2022
Response via : Initial Calibration

