

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042322\
 Data File : VU048254.D
 Acq On : 23 Apr 2022 18:24
 Operator : SY/MD
 Sample : N2485-21
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Apr 25 01:28:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Apr 25 01:24:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	297092	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	295036	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	136373	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	113106	52.500	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	105.000%
7) Chloroethane-d5	1.899	69	86111	50.691	ug/L	0.01
Spiked Amount	50.000	Range	70 - 130	Recovery	=	101.380%
11) 1,1-Dichloroethene-d2	2.565	63	138920	35.244	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	70.480%
21) 2-Butanone-d5	4.623	46	173053	93.584	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	93.580%
24) Chloroform-d	5.063	84	185737	47.118	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.240%
26) 1,2-Dichloroethane-d4	5.703	65	118930	49.191	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.380%
32) Benzene-d6	5.729	84	393007	52.938	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.880%
36) 1,2-Dichloropropane-d6	6.690	67	128606	56.407	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	112.820%
41) Toluene-d8	7.899	98	322496	44.586	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.180%
43) trans-1,3-Dichloroprop...	8.179	79	45515	39.404	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	78.800%
47) 2-Hexanone-d5	8.635	63	121973	91.387	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	91.390%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	206214	52.702	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	105.400%
66) 1,2-Dichlorobenzene-d4	12.195	152	127260	52.452	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	104.900%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042322\
Data File : VU048254.D
Acq On : 23 Apr 2022 18:24
Operator : SY/MD
Sample : N2485-21
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Apr 25 01:28:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Apr 25 01:24:32 2022
Response via : Initial Calibration

