

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042322\
 Data File : VU048255.D
 Acq On : 23 Apr 2022 18:48
 Operator : SY/MD
 Sample : N2486-21
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Apr 25 01:28:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Apr 25 01:24:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	301121	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	305574	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	138408	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	114027	52.219	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 104.440%		
7) Chloroethane-d5	1.893	69	84951	49.339	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	= 98.680%		
11) 1,1-Dichloroethene-d2	2.562	63	139621	34.948	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 69.900%		
21) 2-Butanone-d5	4.626	46	182708	97.483	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 97.480%		
24) Chloroform-d	5.064	84	191682	47.975	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 95.960%		
26) 1,2-Dichloroethane-d4	5.703	65	118647	48.418	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 96.840%		
32) Benzene-d6	5.726	84	400831	52.130	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 104.260%		
36) 1,2-Dichloropropane-d6	6.690	67	133324	56.460	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 112.920%		
41) Toluene-d8	7.899	98	328394	43.836	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 87.680%		
43) trans-1,3-Dichloroprop...	8.179	79	45361	37.917	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 75.840%		
47) 2-Hexanone-d5	8.639	63	128315	92.823	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 92.820%		
56) 1,1,2,2-Tetrachloroeth...	10.758	84	216403	53.399	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 106.800%		
66) 1,2-Dichlorobenzene-d4	12.195	152	133884	54.371	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 108.740%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU042322\
Data File : VU048255.D
Acq On : 23 Apr 2022 18:48
Operator : SY/MD
Sample : N2486-21
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Apr 25 01:28:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Apr 25 01:24:32 2022
Response via : Initial Calibration

