

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU071420\
 Data File : VU039478.D
 Acq On : 14 Jul 2020 12:44
 Operator : SY/MD
 Sample : L3295-02
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BG099

Quant Time: Jul 15 06:33:09 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR070720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Jul 15 06:21:40 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	104168	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	96180	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	46387	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	25328	4.20	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	84.00%
7) Chloroethane-d5	1.92	69	28202	4.57	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	91.40%
11) 1,1-Dichloroethene-d2	2.58	63	40329	3.40	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	68.00%
20) 2-Butanone-d5	4.65	46	119540	43.27	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	86.54%
24) Chloroform-d	5.08	84	58207	4.53	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.60%
26) 1,2-Dichloroethane-d4	5.72	65	36735	4.55	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.00%
32) Benzene-d6	5.74	84	111865	4.83	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.60%
36) 1,2-Dichloropropane-d6	6.70	67	37254	4.58	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	91.60%
41) Toluene-d8	7.91	98	98404	4.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.00%
43) trans-1,3-Dichloropropene-	8.19	79	16140	4.49	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.80%
46) 2-Hexanone-d5	8.64	63	89150	41.84	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	83.68%
57) 1,1,2,2-Tetrachloroethane-	10.76	84	32790	4.07	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	81.40%
64) 1,2-Dichlorobenzene-d4	12.19	152	37657	4.55	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	91.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU071420\
Data File : VU039478.D
Acq On : 14 Jul 2020 12:44
Operator : SY/MD
Sample : L3295-02
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BG099

Quant Time: Jul 15 06:33:09 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR070720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Jul 15 06:21:40 2020
Response via : Initial Calibration

