

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU072120\
 Data File : VU039596.D
 Acq On : 21 Jul 2020 15:01
 Operator : SY/MD
 Sample : L3347-18
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GAY10

Quant Time: Jul 22 02:37:22 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR072120WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Jul 22 02:01:33 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	113618	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	109263	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	49458	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	19879	3.77	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	75.40%
7) Chloroethane-d5	1.92	69	22527	3.91	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	78.20%
11) 1,1-Dichloroethene-d2	2.58	63	34144	2.88	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	57.60%#
20) 2-Butanone-d5	4.65	46	124311	52.26	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.52%
24) Chloroform-d	5.08	84	60453	4.52	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.40%
26) 1,2-Dichloroethane-d4	5.72	65	38466	4.35	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.00%
32) Benzene-d6	5.74	84	104377	3.97	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	79.40%
36) 1,2-Dichloropropane-d6	6.70	67	36924	4.27	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	85.40%
41) Toluene-d8	7.91	98	89185	3.58	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	71.60%
43) trans-1,3-Dichloropropene-	8.19	79	13189	3.51	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	70.20%
46) 2-Hexanone-d5	8.64	63	84336	45.89	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	91.78%
57) 1,1,2,2-Tetrachloroethane-	10.76	84	33895	4.22	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	84.40%
64) 1,2-Dichlorobenzene-d4	12.19	152	37765	4.25	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	85.00%

Target Compounds					Qvalue
51) Chlorobenzene	9.45	112	4499	0.245 ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU072120\
Data File : VU039596.D
Acq On : 21 Jul 2020 15:01
Operator : SY/MD
Sample : L3347-18
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GAY10

Quant Time: Jul 22 02:37:22 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR072120WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Jul 22 02:01:33 2020
Response via : Initial Calibration

