

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090822\
 Data File : VU050694.D
 Acq On : 08 Sep 2022 18:51
 Operator : SY/MD
 Sample : N4483-11ME
 Misc : 4.17g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EW359ME

Quant Time: Sep 08 22:57:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM090222WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Sep 08 22:51:50 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Krupa Patel 09/09/2022
 Supervised By :Mahesh Dadoda 09/09/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.243	114	417838	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	411239	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.816	152	172657	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.594	65	145011m	40.576	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	81.160%		
7) Chloroethane-d5	1.867	69	75446m	31.218	ug/L	-0.03
Spiked Amount 50.000	Range 70 - 130		Recovery =	62.440%#		
11) 1,1-Dichloroethene-d2	2.523	63	196669m	31.141	ug/L	-0.04
Spiked Amount 50.000	Range 60 - 125		Recovery =	62.280%		
21) 2-Butanone-d5	4.633	46	269338	103.009	ug/L	0.01
Spiked Amount 100.000	Range 40 - 130		Recovery =	103.010%		
24) Chloroform-d	5.057	84	308649	45.568	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	91.140%		
26) 1,2-Dichloroethane-d4	5.697	65	199704	48.249	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	96.500%		
32) Benzene-d6	5.716	84	605897	44.583	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	89.160%		
36) 1,2-Dichloropropane-d6	6.687	67	200928	45.032	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	90.060%		
41) Toluene-d8	7.896	98	525326	45.202	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	90.400%		
43) trans-1,3-Dichloroprop...	8.182	79	88371	47.868	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	95.740%		
47) 2-Hexanone-d5	8.655	63	170242	151.685	ug/L	0.02
Spiked Amount 100.000	Range 45 - 130		Recovery =	151.690%#		
56) 1,1,2,2-Tetrachloroeth...	10.764	84	335352	49.666	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	99.340%		
66) 1,2-Dichlorobenzene-d4	12.195	152	200686	49.653	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	99.300%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090822\
Data File : VU050694.D
Acq On : 08 Sep 2022 18:51
Operator : SY/MD
Sample : N4483-11ME
Misc : 4.17g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EW359ME

Quant Time: Sep 08 22:57:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM090222WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Sep 08 22:51:50 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 09/09/2022
Supervised By :Mahesh Dadoda 09/09/2022

