

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110122\
 Data File : VU051658.D
 Acq On : 01 Nov 2022 16:51
 Operator : JC/MD
 Sample : N5276-03
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Nov 02 02:04:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Nov 02 01:48:00 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	202100	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	211235	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	89095	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	82104	5.394	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	107.800%	
7) Chloroethane-d5	1.909	69	72375	4.559	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	91.200%	
11) 1,1-Dichloroethene-d2	2.565	65	33172	6.005	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	120.000%	
20) 2-Butanone-d5	4.620	46	182102	33.035	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	66.060%	
24) Chloroform-d	5.060	84	154507	4.338	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	86.800%	
26) 1,2-Dichloroethane-d4	5.700	65	77236	4.014	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	80.200%	
32) Benzene-d6	5.726	84	307465	4.612	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	92.200%	
36) 1,2-Dichloropropane-d6	6.687	67	101734	4.365	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	87.200%	
41) Toluene-d8	7.896	98	250110	4.771	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	95.400%	
43) trans-1,3-Dichloroprop...	8.179	79	31461	4.414	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	88.200%	
46) 2-Hexanone-d5	8.629	63	148782	34.678	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	69.360%	
56) 1,1,2,2-Tetrachloroeth...	10.755	84	83691	3.918	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	78.400%	
66) 1,2-Dichlorobenzene-d4	12.192	152	87484	5.350	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	107.000%	
Target Compounds						Qvalue
16) Methylene chloride	3.044	84	7214	0.251	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110122\
Data File : VU051658.D
Acq On : 01 Nov 2022 16:51
Operator : JC/MD
Sample : N5276-03
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Nov 02 02:04:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 02 01:48:00 2022
Response via : Initial Calibration

