

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU110725\  
 Data File : VU063678.D  
 Acq On : 07 Nov 2025 08:55  
 Operator : MD/SY  
 Sample : VSTDCCC010  
 Misc : 25.0mL/MSVOA\_U/WATER  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 LabSampleId :  
 VSTDCCC010

Quant Time: Nov 08 02:30:48 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\524U103125DW.M  
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER  
 QLast Update : Mon Nov 03 00:51:51 2025  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 30% Max. Rel. Area : 200%

|       | Compound                    | AvgRF | CCRF  | %Dev  | Area% | Dev(min) |
|-------|-----------------------------|-------|-------|-------|-------|----------|
| 1 i   | Fluorobenzene               | 1.000 | 1.000 | 0.0   | 93    | 0.00     |
| 2 T   | Dichlorodifluoromethane     | 0.376 | 0.391 | -4.0  | 92    | 0.00     |
| 3 t   | Chloromethane               | 0.354 | 0.349 | 1.4   | 92    | 0.00     |
| 4 Rt  | Vinyl Chloride              | 0.360 | 0.383 | -6.4  | 93    | 0.00     |
| 5 T   | Bromomethane                | 0.218 | 0.223 | -2.3  | 99    | 0.01     |
| 6 T   | Chloroethane                | 0.221 | 0.239 | -8.1  | 96    | 0.00     |
| 7 T   | Trichlorofluoromethane      | 0.524 | 0.576 | -9.9  | 96    | 0.00     |
| 8     | 1,1,2-Trichloro-1,2,2-trifl | 0.278 | 0.313 | -12.6 | 99    | 0.00     |
| 9 Rt  | 1,1-Dichloroethene          | 0.270 | 0.294 | -8.9  | 96    | 0.00     |
| 10 t  | Iodomethane                 | 0.400 | 0.442 | -10.5 | 94    | 0.00     |
| 11 t  | Allyl Chloride              | 0.421 | 0.442 | -5.0  | 92    | 0.00     |
| 12 t  | Acrylonitrile               | 0.067 | 0.066 | 1.5   | 83    | 0.00     |
| 13 T  | Acetone                     | 0.118 | 0.139 | -17.8 | 109   | 0.00     |
| 14 T  | Carbon Disulfide            | 0.760 | 0.776 | -2.1  | 92    | 0.00     |
| 15 RT | Methylene Chloride          | 0.347 | 0.333 | 4.0   | 95    | 0.00     |
| 16 RT | trans-1,2-Dichloroethene    | 0.297 | 0.322 | -8.4  | 94    | 0.00     |
| 17 t  | 1,1-Dichloroethane          | 0.589 | 0.637 | -8.1  | 96    | 0.00     |
| 18 T  | 2-Butanone                  | 0.125 | 0.142 | -13.6 | 101   | 0.00     |
| 19    | Cyclohexane                 | 0.450 | 0.486 | -8.0  | 93    | 0.00     |
| 20    | Methylcyclohexane           | 0.560 | 0.565 | -0.9  | 88    | 0.00     |
| 21 T  | 2,2-Dichloropropane         | 0.528 | 0.565 | -7.0  | 95    | 0.00     |
| 22 RT | cis-1,2-Dichloroethene      | 0.339 | 0.370 | -9.1  | 95    | 0.00     |
| 23 t  | Diethyl Ether               | 0.221 | 0.227 | -2.7  | 88    | 0.00     |
| 24 t  | tert-Butyl Alcohol          | 0.038 | 0.026 | 31.6# | 66    | -0.01    |
| 25 t  | Methyl tert-Butyl Ether     | 0.790 | 0.788 | 0.3   | 87    | 0.00     |
| 26 t  | Bromochloromethane          | 0.143 | 0.149 | -4.2  | 92    | 0.00     |
| 27 t  | Chloroform                  | 0.603 | 0.658 | -9.1  | 97    | 0.00     |
| 28 RT | 1,1,1-Trichloroethane       | 0.531 | 0.575 | -8.3  | 95    | 0.00     |
| 29 T  | 1,1-Dichloropropene         | 0.446 | 0.494 | -10.8 | 96    | 0.00     |
| 30 RT | Carbon Tetrachloride        | 0.421 | 0.461 | -9.5  | 93    | 0.00     |
| 31 t  | Isopropyl Ether             | 0.933 | 1.001 | -7.3  | 94    | 0.00     |
| 32    | Ethyl-t-butyl ether         | 0.899 | 0.899 | 0.0   | 89    | 0.00     |
| 33    | Tert-Amyl methyl ether      | 0.783 | 0.788 | -0.6  | 88    | 0.00     |
| 34 t  | Propionitrile               | 0.024 | 0.023 | 4.2   | 78    | 0.00     |
| 35 RT | Benzene                     | 1.257 | 1.367 | -8.8  | 96    | 0.00     |
| 36 RT | 1,2-Dichloroethane          | 0.404 | 0.427 | -5.7  | 91    | 0.00     |
| 37 RT | Trichloroethene             | 0.358 | 0.372 | -3.9  | 97    | 0.00     |
| 38 Rt | 1,2-Dichloropropane         | 0.352 | 0.356 | -1.1  | 89    | 0.00     |
| 39 t  | Methacrylonitrile           | 0.122 | 0.105 | 13.9  | 82    | 0.00     |
| 40 t  | Methyl acrylate             | 0.177 | 0.155 | 12.4  | 82    | 0.00     |
| 41 t  | Tetrahydrofuran             | 0.055 | 0.053 | 3.6   | 80    | 0.00     |
| 42 t  | 1-Chlorobutane              | 0.598 | 0.673 | -12.5 | 94    | 0.00     |
| 43 T  | Dibromomethane              | 0.182 | 0.177 | 2.7   | 83    | 0.00     |
| 44 T  | Bromodichloromethane        | 0.445 | 0.444 | 0.2   | 86    | 0.00     |
| 45 T  | 4-Methyl-2-Pentanone        | 0.220 | 0.203 | 7.7   | 78    | 0.00     |
| 46 t  | t-1,4-Dichloro-2-butene     | 0.076 | 0.077 | -1.3  | 84    | 0.00     |
| 47 t  | Methyl methacrylate         | 0.167 | 0.157 | 6.0   | 77    | 0.00     |
| 48 t  | Ethyl methacrylate          | 0.341 | 0.321 | 5.9   | 77    | 0.00     |
| 49 Rt | Toluene                     | 0.848 | 0.888 | -4.7  | 91    | 0.00     |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU110725\  
 Data File : VU063678.D  
 Acq On : 07 Nov 2025 08:55  
 Operator : MD/SY  
 Sample : VSTDCCC010  
 Misc : 25.0mL/MSVOA\_U/WATER  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 LabSampleId :  
 VSTDCCC010

Quant Time: Nov 08 02:30:48 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\524U103125DW.M  
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER  
 QLast Update : Mon Nov 03 00:51:51 2025  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 30% Max. Rel. Area : 200%

|       | Compound                    | AvgRF | CCRF  | %Dev  | Area% | Dev(min) |
|-------|-----------------------------|-------|-------|-------|-------|----------|
| 50 T  | t-1,3-Dichloropropene       | 0.387 | 0.350 | 9.6   | 79    | 0.00     |
| 51 T  | cis-1,3-Dichloropropene     | 0.484 | 0.477 | 1.4   | 85    | 0.00     |
| 52 RT | 1,1,2-Trichloroethane       | 0.247 | 0.253 | -2.4  | 87    | 0.00     |
| 53 t  | 1,3-Dichloropropane         | 0.442 | 0.433 | 2.0   | 86    | 0.00     |
| 54 t  | 2-Hexanone                  | 0.185 | 0.207 | -11.9 | 95    | 0.00     |
| 55 t  | Dibromochloromethane        | 0.291 | 0.282 | 3.1   | 82    | 0.00     |
| 56 T  | 1,2-Dibromoethane           | 0.236 | 0.225 | 4.7   | 84    | 0.00     |
| 57 S  | 4-Bromofluorobenzene        | 0.452 | 0.546 | -20.8 | 106   | 0.00     |
| 58 RT | Tetrachloroethene           | 0.351 | 0.392 | -11.7 | 95    | 0.00     |
| 59 Rt | Chlorobenzene               | 0.920 | 0.971 | -5.5  | 93    | 0.00     |
| 60 T  | 1,1,1,2-Tetrachloroethane   | 0.322 | 0.320 | 0.6   | 88    | 0.00     |
| 61 t  | Pentachloroethane           | 0.223 | 0.207 | 7.2   | 82    | 0.00     |
| 62 t  | Hexachloroethane            | 0.245 | 0.241 | 1.6   | 83    | 0.00     |
| 63 Rt | Ethyl Benzene               | 1.666 | 1.732 | -4.0  | 92    | 0.00     |
| 64 RT | m/p-Xylenes                 | 0.622 | 0.651 | -4.7  | 91    | 0.00     |
| 65 RT | o-Xylene                    | 0.609 | 0.637 | -4.6  | 90    | 0.00     |
| 66 RT | Styrene                     | 0.989 | 1.030 | -4.1  | 89    | 0.00     |
| 67 t  | Bromoform                   | 0.157 | 0.140 | 10.8  | 75    | 0.00     |
| 68 S  | 1,2-Dichlorobenzene-d4      | 0.406 | 0.395 | 2.7   | 81    | 0.00     |
| 69 T  | Isopropylbenzene            | 1.505 | 1.573 | -4.5  | 91    | 0.00     |
| 70 T  | 1,1,2,2-Tetrachloroethane   | 0.311 | 0.293 | 5.8   | 80    | 0.00     |
| 71 T  | 1,2,3-Trichloropropane      | 0.270 | 0.237 | 12.2  | 81    | 0.00     |
| 72 t  | Bromobenzene                | 0.364 | 0.377 | -3.6  | 90    | 0.00     |
| 73 t  | n-propylbenzene             | 0.428 | 0.455 | -6.3  | 91    | 0.00     |
| 74 t  | 2-Chlorotoluene             | 0.371 | 0.394 | -6.2  | 93    | 0.00     |
| 75 t  | 1,3,5-Trimethylbenzene      | 1.392 | 1.447 | -4.0  | 90    | 0.00     |
| 76 t  | 4-Chlorotoluene             | 0.372 | 0.400 | -7.5  | 93    | 0.00     |
| 77 t  | tert-Butylbenzene           | 1.363 | 1.373 | -0.7  | 88    | 0.00     |
| 78 t  | 1,2,4-Trimethylbenzene      | 1.336 | 1.427 | -6.8  | 90    | 0.00     |
| 79 t  | sec-Butylbenzene            | 1.793 | 1.863 | -3.9  | 89    | 0.00     |
| 80    | Nitrobenzene                | 0.011 | 0.005 | 54.5# | 38    | 0.00     |
| 81 t  | p-Isopropyltoluene          | 1.469 | 1.531 | -4.2  | 89    | 0.00     |
| 82 t  | 1,3-Dichlorobenzene         | 0.705 | 0.751 | -6.5  | 91    | 0.00     |
| 83 Rt | 1,4-Dichlorobenzene         | 0.700 | 0.736 | -5.1  | 90    | 0.00     |
| 84 t  | n-Butylbenzene              | 1.316 | 1.389 | -5.5  | 87    | 0.00     |
| 85 Rt | 1,2-Dichlorobenzene         | 0.667 | 0.693 | -3.9  | 89    | 0.00     |
| 86 t  | 1,2-Dibromo-3-Chloropropane | 0.052 | 0.045 | 13.5  | 70    | 0.00     |
| 87 Rt | 1,2,4-Trichlorobenzene      | 0.411 | 0.419 | -1.9  | 83    | 0.00     |
| 88 t  | Hexachlorobutadiene         | 0.245 | 0.270 | -10.2 | 94    | 0.00     |
| 89 t  | Naphthalene                 | 0.790 | 0.654 | 17.2  | 69    | 0.00     |
| 90 t  | 1,2,3-Trichlorobenzene      | 0.375 | 0.367 | 2.1   | 79    | 0.00     |

(#) = Out of Range

SPCC's out = 0 CCC's out = 0