

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU123021\
 Data File : VU046651.D
 Acq On : 30 Dec 2021 16:12
 Operator : SY/MD
 Sample : M5221-02
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PW-B6-L10-122221-FD

Quant Time: Dec 31 07:26:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U123021DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Thu Dec 30 12:44:59 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.118	96	16688	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.636	95	4621	0.778	ug/l	0.00
Spiked Amount 1.000			Recovery	=	78.000%	
68) 1,2-Dichlorobenzene-d4	12.195	152	5293	0.824	ug/l	0.00
Spiked Amount 1.000			Recovery	=	82.000%	
Target Compounds						
						Qvalue
8) 1,1,2-Trichloro-1,2,2-...	2.584	101	1411	0.298	ug/l	90
15) Methylene Chloride	3.051	84	1912	0.194	ug/l	91
25) Methyl tert-Butyl Ether	3.366	73	2029	0.189	ug/l	# 74
37) Trichloroethene	6.546	130	4483	0.940	ug/l	95
58) Tetrachloroethene	8.555	164	6917	1.640	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU123021\
Data File : VU046651.D
Acq On : 30 Dec 2021 16:12
Operator : SY/MD
Sample : M5221-02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PW-B6-L10-122221-FD

Quant Time: Dec 31 07:26:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U123021DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Thu Dec 30 12:44:59 2021
Response via : Initial Calibration

