




Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011922\
 Data File : VU046794.D
 Acq On : 19 Jan 2022 12:57
 Operator : SY/MD
 Sample : N1179-05
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0HJ8

Manual IntegrationsAPPROVED

Quant Time: Jan 20 02:01:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jan 20 02:00:32 2022
 Response via : Initial Calibration

Reviewed By :John Carlone 01/20/2022
 Supervised By :Mahesh Dadoda 01/21/2022



TIC: VU046794.D\data.ms

(13) Acetone (T)

2.678min (-0.013) 3.64 ug/L

response 5321

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	35.80	15.39
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011922\
Data File : VU046794.D
Acq On : 19 Jan 2022 12:57
Operator : SY/MD
Sample : N1179-05
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0HJ8

Manual IntegrationsAPPROVED

Quant Time: Jan 20 02:01:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jan 20 02:00:32 2022
Response via : Initial Calibration

Reviewed By :John Carlone 01/20/2022
Supervised By :Mahesh Dadoda 01/21/2022



TIC: VU046794.D\data.ms

(13) Acetone (T)

2.678min (-0.013) 7.28 ug/L m

response 10653

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	35.80	7.69
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011922\
 Data File : VU046794.D
 Acq On : 19 Jan 2022 12:57
 Operator : SY/MD
 Sample : N1179-05
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 COHJ8

Manual IntegrationsAPPROVED

Quant Time: Jan 20 02:01:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jan 20 02:00:32 2022
 Response via : Initial Calibration

Reviewed By : John Carlone 01/20/2022
 Supervised By : Mahesh Dadoda 01/21/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	107109	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	114204	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.816	152	58022	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	46061	5.075	ug/L	0.00
Spiked Amount 5.000	Range 40 - 130		Recovery	=	101.600%	
7) Chloroethane-d5	1.919	69	38636	5.403	ug/L	0.00
Spiked Amount 5.000	Range 65 - 130		Recovery	=	108.000%	
11) 1,1-Dichloroethene-d2	2.572	65	18637	5.162	ug/L	0.00
Spiked Amount 5.000	Range 60 - 125		Recovery	=	103.200%	
20) 2-Butanone-d5	4.655	46	157864	58.713	ug/L	0.00
Spiked Amount 50.000	Range 40 - 130		Recovery	=	117.420%	
24) Chloroform-d	5.070	84	82872	5.201	ug/L	0.00
Spiked Amount 5.000	Range 70 - 125		Recovery	=	104.000%	
26) 1,2-Dichloroethane-d4	5.707	65	47802	5.509	ug/L	0.00
Spiked Amount 5.000	Range 70 - 130		Recovery	=	110.200%	
32) Benzene-d6	5.732	84	173599	5.172	ug/L	0.00
Spiked Amount 5.000	Range 70 - 125		Recovery	=	103.400%	
36) 1,2-Dichloropropane-d6	6.694	67	52707	5.230	ug/L	0.00
Spiked Amount 5.000	Range 60 - 140		Recovery	=	104.600%	
41) Toluene-d8	7.903	98	148099	4.821	ug/L	0.00
Spiked Amount 5.000	Range 70 - 130		Recovery	=	96.400%	
43) trans-1,3-Dichloroprop...	8.186	79	21144	4.777	ug/L	0.00
Spiked Amount 5.000	Range 55 - 130		Recovery	=	95.600%	
46) 2-Hexanone-d5	8.639	63	139335	53.169	ug/L	0.00
Spiked Amount 50.000	Range 45 - 130		Recovery	=	106.340%	
56) 1,1,2,2-Tetrachloroeth...	10.761	84	52731	5.212	ug/L	0.00
Spiked Amount 5.000	Range 65 - 120		Recovery	=	104.200%	
66) 1,2-Dichlorobenzene-d4	12.195	152	58858	5.409	ug/L	0.00
Spiked Amount 5.000	Range 80 - 120		Recovery	=	108.200%	
Target Compounds						Qvalue
3) Chloromethane	1.520	50	1078	0.130	ug/L	97
13) Acetone	2.678	43	10653m	7.285	ug/L	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011922\
Data File : VU046794.D
Acq On : 19 Jan 2022 12:57
Operator : SY/MD
Sample : N1179-05
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0HJ8

Manual IntegrationsAPPROVED

Quant Time: Jan 20 02:01:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jan 20 02:00:32 2022
Response via : Initial Calibration

Reviewed By :John Carlone 01/20/2022
Supervised By :Mahesh Dadoda 01/21/2022

