

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU012622\  
 Data File : VU046840.D  
 Acq On : 26 Jan 2022 12:57  
 Operator : SY/MD  
 Sample : N1260-01 20X  
 Misc : 25.0mL/MSVOA\_U/WATER  
 ALS Vial : 6 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampleId :  
 BFXX3

Quant Time: Jan 26 14:51:40 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\SFAMUTR011822WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Mon Jan 24 00:23:26 2022  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon  | Response | Conc     | Units | Dev(Min) |
|-------------------------------|--------|-------|----------|----------|-------|----------|
| -----                         |        |       |          |          |       |          |
| Internal Standards            |        |       |          |          |       |          |
| 1) 1,4-Difluorobenzene        | 6.253  | 114   | 99022    | 5.000    | ug/L  | 0.00     |
| 28) Chlorobenzene-d5          | 9.420  | 117   | 110727   | 5.000    | ug/L  | 0.00     |
| 58) 1,4-Dichlorobenzene-d4    | 11.812 | 152   | 49592    | 5.000    | ug/L  | 0.00     |
|                               |        |       |          |          |       |          |
| System Monitoring Compounds   |        |       |          |          |       |          |
| 4) Vinyl Chloride-d3          | 1.604  | 65    | 38614    | 4.602    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 40 - 130 | Recovery | =     | 92.000%  |
| 7) Chloroethane-d5            | 1.919  | 69    | 33567    | 5.077    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 65 - 130 | Recovery | =     | 101.600% |
| 11) 1,1-Dichloroethene-d2     | 2.575  | 65    | 16865    | 5.053    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 60 - 125 | Recovery | =     | 101.000% |
| 20) 2-Butanone-d5             | 4.652  | 46    | 109302   | 43.971   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000 | Range | 40 - 130 | Recovery | =     | 87.940%  |
| 24) Chloroform-d              | 5.070  | 84    | 75416    | 5.120    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 125 | Recovery | =     | 102.400% |
| 26) 1,2-Dichloroethane-d4     | 5.706  | 65    | 43428    | 5.414    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 130 | Recovery | =     | 108.200% |
| 32) Benzene-d6                | 5.732  | 84    | 146049   | 4.488    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 125 | Recovery | =     | 89.800%  |
| 36) 1,2-Dichloropropane-d6    | 6.693  | 67    | 46422    | 4.751    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 60 - 140 | Recovery | =     | 95.000%  |
| 41) Toluene-d8                | 7.899  | 98    | 118582   | 3.981    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 130 | Recovery | =     | 79.600%  |
| 43) trans-1,3-Dichloroprop... | 8.182  | 79    | 16683    | 3.888    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 55 - 130 | Recovery | =     | 77.800%  |
| 46) 2-Hexanone-d5             | 8.639  | 63    | 101066   | 39.777   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000 | Range | 45 - 130 | Recovery | =     | 79.560%  |
| 56) 1,1,2,2-Tetrachloroeth... | 10.758 | 84    | 42897    | 4.373    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 65 - 120 | Recovery | =     | 87.400%  |
| 66) 1,2-Dichlorobenzene-d4    | 12.195 | 152   | 44786    | 4.816    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 80 - 120 | Recovery | =     | 96.400%  |
|                               |        |       |          |          |       |          |
| Target Compounds              |        |       |          |          |       | Qvalue   |
| 38) Bromodichloromethane      | 7.108  | 83    | 438      | 0.048    | ug/L  | # 96     |
| -----                         |        |       |          |          |       |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU012622\  
Data File : VU046840.D  
Acq On : 26 Jan 2022 12:57  
Operator : SY/MD  
Sample : N1260-01 20X  
Misc : 25.0mL/MSVOA\_U/WATER  
ALS Vial : 6 Sample Multiplier: 1

Instrument :  
MSVOA\_U  
ClientSampleId :  
BFXX3

Quant Time: Jan 26 14:51:40 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\SFAMUTR011822WMA.M  
Quant Title : TRACE VOA SFAM1.0  
QLast Update : Mon Jan 24 00:23:26 2022  
Response via : Initial Calibration

