

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU012622\  
 Data File : VU046854.D  
 Acq On : 26 Jan 2022 19:05  
 Operator : SY/MD  
 Sample : N1277-04  
 Misc : 25.0mL/MSVOA\_U/WATER  
 ALS Vial : 21 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampleId :  
 C0VD8

Quant Time: Jan 27 00:16:00 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\SFAMUTR011822WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Thu Jan 27 00:12:19 2022  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon  | Response | Conc     | Units | Dev(Min) |
|-------------------------------|--------|-------|----------|----------|-------|----------|
| -----                         |        |       |          |          |       |          |
| Internal Standards            |        |       |          |          |       |          |
| 1) 1,4-Difluorobenzene        | 6.253  | 114   | 99544    | 5.000    | ug/L  | 0.00     |
| 28) Chlorobenzene-d5          | 9.417  | 117   | 108198   | 5.000    | ug/L  | 0.00     |
| 58) 1,4-Dichlorobenzene-d4    | 11.812 | 152   | 48371    | 5.000    | ug/L  | 0.00     |
| System Monitoring Compounds   |        |       |          |          |       |          |
| 4) Vinyl Chloride-d3          | 1.604  | 65    | 35330    | 4.189    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 40 - 130 | Recovery | =     | 83.800%  |
| 7) Chloroethane-d5            | 1.919  | 69    | 31823    | 4.788    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 65 - 130 | Recovery | =     | 95.800%  |
| 11) 1,1-Dichloroethene-d2     | 2.572  | 65    | 15188    | 4.527    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 60 - 125 | Recovery | =     | 90.600%  |
| 20) 2-Butanone-d5             | 4.655  | 46    | 103060   | 41.243   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000 | Range | 40 - 130 | Recovery | =     | 82.480%  |
| 24) Chloroform-d              | 5.067  | 84    | 72531    | 4.898    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 125 | Recovery | =     | 98.000%  |
| 26) 1,2-Dichloroethane-d4     | 5.706  | 65    | 41236    | 5.114    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 130 | Recovery | =     | 102.200% |
| 32) Benzene-d6                | 5.732  | 84    | 140020   | 4.403    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 125 | Recovery | =     | 88.000%  |
| 36) 1,2-Dichloropropane-d6    | 6.694  | 67    | 44504    | 4.661    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 60 - 140 | Recovery | =     | 93.200%  |
| 41) Toluene-d8                | 7.899  | 98    | 111605   | 3.834    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 130 | Recovery | =     | 76.600%  |
| 43) trans-1,3-Dichloroprop... | 8.182  | 79    | 16351    | 3.900    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 55 - 130 | Recovery | =     | 78.000%  |
| 46) 2-Hexanone-d5             | 8.639  | 63    | 97740    | 39.367   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000 | Range | 45 - 130 | Recovery | =     | 78.740%  |
| 56) 1,1,2,2-Tetrachloroeth... | 10.758 | 84    | 40591    | 4.235    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 65 - 120 | Recovery | =     | 84.600%  |
| 66) 1,2-Dichlorobenzene-d4    | 12.195 | 152   | 42346    | 4.668    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 80 - 120 | Recovery | =     | 93.400%  |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU012622\  
Data File : VU046854.D  
Acq On : 26 Jan 2022 19:05  
Operator : SY/MD  
Sample : N1277-04  
Misc : 25.0mL/MSVOA\_U/WATER  
ALS Vial : 21 Sample Multiplier: 1

Instrument :  
MSVOA\_U  
ClientSampleId :  
C0VD8

Quant Time: Jan 27 00:16:00 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\SFAMUTR011822WMA.M  
Quant Title : TRACE VOA SFAM1.0  
QLast Update : Thu Jan 27 00:12:19 2022  
Response via : Initial Calibration

