

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU012622\  
 Data File : VU046856.D  
 Acq On : 26 Jan 2022 19:54  
 Operator : SY/MD  
 Sample : N1277-06  
 Misc : 25.0mL/MSVOA\_U/WATER  
 ALS Vial : 23 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampleId :  
 COVFO

Quant Time: Jan 27 00:16:20 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\SFAMUTR011822WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Thu Jan 27 00:12:19 2022  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon  | Response | Conc     | Units | Dev(Min)   |
|-------------------------------|--------|-------|----------|----------|-------|------------|
| -----                         |        |       |          |          |       |            |
| Internal Standards            |        |       |          |          |       |            |
| 1) 1,4-Difluorobenzene        | 6.253  | 114   | 98114    | 5.000    | ug/L  | 0.00       |
| 28) Chlorobenzene-d5          | 9.420  | 117   | 105819   | 5.000    | ug/L  | 0.00       |
| 58) 1,4-Dichlorobenzene-d4    | 11.812 | 152   | 47234    | 5.000    | ug/L  | 0.00       |
| System Monitoring Compounds   |        |       |          |          |       |            |
| 4) Vinyl Chloride-d3          | 1.604  | 65    | 36296    | 4.366    | ug/L  | 0.00       |
| Spiked Amount                 | 5.000  | Range | 40 - 130 | Recovery | =     | 87.400%    |
| 7) Chloroethane-d5            | 1.919  | 69    | 31207    | 4.764    | ug/L  | 0.00       |
| Spiked Amount                 | 5.000  | Range | 65 - 130 | Recovery | =     | 95.200%    |
| 11) 1,1-Dichloroethene-d2     | 2.575  | 65    | 15286    | 4.622    | ug/L  | 0.00       |
| Spiked Amount                 | 5.000  | Range | 60 - 125 | Recovery | =     | 92.400%    |
| 20) 2-Butanone-d5             | 4.645  | 46    | 110338   | 44.799   | ug/L  | 0.00       |
| Spiked Amount                 | 50.000 | Range | 40 - 130 | Recovery | =     | 89.600%    |
| 24) Chloroform-d              | 5.070  | 84    | 71626    | 4.908    | ug/L  | 0.00       |
| Spiked Amount                 | 5.000  | Range | 70 - 125 | Recovery | =     | 98.200%    |
| 26) 1,2-Dichloroethane-d4     | 5.706  | 65    | 39917    | 5.022    | ug/L  | 0.00       |
| Spiked Amount                 | 5.000  | Range | 70 - 130 | Recovery | =     | 100.400%   |
| 32) Benzene-d6                | 5.732  | 84    | 139119   | 4.473    | ug/L  | 0.00       |
| Spiked Amount                 | 5.000  | Range | 70 - 125 | Recovery | =     | 89.400%    |
| 36) 1,2-Dichloropropane-d6    | 6.693  | 67    | 44028    | 4.715    | ug/L  | 0.00       |
| Spiked Amount                 | 5.000  | Range | 60 - 140 | Recovery | =     | 94.400%    |
| 41) Toluene-d8                | 7.899  | 98    | 110719   | 3.889    | ug/L  | 0.00       |
| Spiked Amount                 | 5.000  | Range | 70 - 130 | Recovery | =     | 77.800%    |
| 43) trans-1,3-Dichloroprop... | 8.182  | 79    | 16338    | 3.984    | ug/L  | 0.00       |
| Spiked Amount                 | 5.000  | Range | 55 - 130 | Recovery | =     | 79.600%    |
| 46) 2-Hexanone-d5             | 8.636  | 63    | 94445    | 38.895   | ug/L  | 0.00       |
| Spiked Amount                 | 50.000 | Range | 45 - 130 | Recovery | =     | 77.800%    |
| 56) 1,1,2,2-Tetrachloroeth... | 10.758 | 84    | 39876    | 4.253    | ug/L  | 0.00       |
| Spiked Amount                 | 5.000  | Range | 65 - 120 | Recovery | =     | 85.000%    |
| 66) 1,2-Dichlorobenzene-d4    | 12.195 | 152   | 42873    | 4.840    | ug/L  | 0.00       |
| Spiked Amount                 | 5.000  | Range | 80 - 120 | Recovery | =     | 96.800%    |
| Target Compounds              |        |       |          |          |       |            |
| 33) Benzene                   | 5.784  | 78    | 2225     | 0.089    | ug/L  | Qvalue 100 |
| -----                         |        |       |          |          |       |            |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU012622\  
Data File : VU046856.D  
Acq On : 26 Jan 2022 19:54  
Operator : SY/MD  
Sample : N1277-06  
Misc : 25.0mL/MSVOA\_U/WATER  
ALS Vial : 23 Sample Multiplier: 1

Instrument :  
MSVOA\_U  
ClientSampleId :  
C0VF0

Quant Time: Jan 27 00:16:20 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\SFAMUTR011822WMA.M  
Quant Title : TRACE VOA SFAM1.0  
QLast Update : Thu Jan 27 00:12:19 2022  
Response via : Initial Calibration

