

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU012622\  
 Data File : VU046838.D  
 Acq On : 26 Jan 2022 12:03  
 Operator : SY/MD  
 Sample : VU0126WBL01  
 Misc : 25.0mL/MSVOA\_U/WATER  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampleId :  
 VBLK054

Quant Time: Jan 26 14:51:27 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\SFAMUTR011822WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Mon Jan 24 00:23:26 2022  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	96252	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	105578	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	46303	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	37881	4.645	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	92.800%	
7) Chloroethane-d5	1.919	69	33228	5.171	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	103.400%	
11) 1,1-Dichloroethene-d2	2.575	65	16747	5.162	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	103.200%	
20) 2-Butanone-d5	4.658	46	107845	44.634	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	89.260%	
24) Chloroform-d	5.070	84	75363	5.264	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	105.200%	
26) 1,2-Dichloroethane-d4	5.707	65	42454	5.445	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	108.800%	
32) Benzene-d6	5.732	84	145054	4.675	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	93.400%	
36) 1,2-Dichloropropane-d6	6.694	67	44932	4.823	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	96.400%	
41) Toluene-d8	7.899	98	116085	4.087	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	81.800%	
43) trans-1,3-Dichloroprop...	8.182	79	16971	4.148	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	83.000%	
46) 2-Hexanone-d5	8.639	63	100744	41.584	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	83.160%	
56) 1,1,2,2-Tetrachloroeth...	10.758	84	42959	4.593	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	91.800%	
66) 1,2-Dichlorobenzene-d4	12.195	152	43742	5.037	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	100.800%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU012622\  
 Data File : VU046838.D  
 Acq On : 26 Jan 2022 12:03  
 Operator : SY/MD  
 Sample : VU0126WBL01  
 Misc : 25.0mL/MSVOA\_U/WATER  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampleId :  
 VBLK054

Quant Time: Jan 26 14:51:27 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\SFAMUTR011822WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Mon Jan 24 00:23:26 2022  
 Response via : Initial Calibration

