

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012822\
 Data File : VU046891.D
 Acq On : 28 Jan 2022 14:31
 Operator : SY/MD
 Sample : VIBLK025
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK025

Quant Time: Jan 29 00:40:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Jan 29 00:38:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	95119	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	106080	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	45327	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	32008	3.971	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	79.400%
7) Chloroethane-d5	1.919	69	29829	4.697	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.000%
11) 1,1-Dichloroethene-d2	2.572	65	14599	4.554	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	91.000%
20) 2-Butanone-d5	4.645	46	101092	42.337	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	84.680%
24) Chloroform-d	5.067	84	73535	5.197	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.000%
26) 1,2-Dichloroethane-d4	5.706	65	41076	5.331	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.600%
32) Benzene-d6	5.732	84	137007	4.394	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.800%
36) 1,2-Dichloropropane-d6	6.694	67	45747	4.887	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.800%
41) Toluene-d8	7.899	98	105001	3.680	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	73.600%
43) trans-1,3-Dichloroprop...	8.182	79	15698	3.819	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	76.400%
46) 2-Hexanone-d5	8.636	63	100663	41.354	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	82.700%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	43251	4.602	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.000%
66) 1,2-Dichlorobenzene-d4	12.195	152	40288	4.740	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	94.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012822\
Data File : VU046891.D
Acq On : 28 Jan 2022 14:31
Operator : SY/MD
Sample : VIBLK025
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK025

Quant Time: Jan 29 00:40:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jan 29 00:38:36 2022
Response via : Initial Calibration

