

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012822\
 Data File : VU046892.D
 Acq On : 28 Jan 2022 14:56
 Operator : SY/MD
 Sample : N1273-06DL 40X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0HK5DL

Quant Time: Jan 29 00:41:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Jan 29 00:38:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	94584	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	102046	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	44036	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	31005	3.869	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	77.400%
7) Chloroethane-d5	1.919	69	30736	4.867	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	97.400%
11) 1,1-Dichloroethene-d2	2.575	65	14163	4.442	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	88.800%
20) 2-Butanone-d5	4.649	46	123137	51.861	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.720%
24) Chloroform-d	5.067	84	72119	5.126	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.600%
26) 1,2-Dichloroethane-d4	5.707	65	40130	5.238	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.800%
32) Benzene-d6	5.729	84	135868	4.530	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.600%
36) 1,2-Dichloropropane-d6	6.694	67	44831	4.979	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.600%
41) Toluene-d8	7.903	98	102429	3.731	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	74.600%
43) trans-1,3-Dichloroprop...	8.179	79	15663	3.961	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	79.200%
46) 2-Hexanone-d5	8.636	63	102646	43.836	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	87.680%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	43162	4.774	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.400%
66) 1,2-Dichlorobenzene-d4	12.195	152	38877	4.708	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	94.200%
Target Compounds						Qvalue
5) Vinyl chloride	1.607	62	23849	3.127	ug/L	98
12) 1,1-Dichloroethene	2.578	96	979	0.185	ug/L	# 1
19) 1,1-Dichloroethane	3.877	63	3719	0.387	ug/L	97
22) cis-1,2-Dichloroethene	4.678	96	2339	0.388	ug/L	93
34) Trichloroethene	6.546	95	6592	1.057	ug/L	91
47) Tetrachloroethene	8.555	164	40793	8.041	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012822\
Data File : VU046892.D
Acq On : 28 Jan 2022 14:56
Operator : SY/MD
Sample : N1273-06DL 40X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0HK5DL

Quant Time: Jan 29 00:41:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jan 29 00:38:36 2022
Response via : Initial Calibration

