

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012822\
 Data File : VU046894.D
 Acq On : 28 Jan 2022 15:45
 Operator : SY/MD
 Sample : VIBLK026
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK026

Quant Time: Jan 29 00:41:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Jan 29 00:38:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	97734	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	108715	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	46440	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	29768	3.594	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	71.800%
7) Chloroethane-d5	1.919	69	29264	4.485	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.600%
11) 1,1-Dichloroethene-d2	2.572	65	13147	3.991	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	79.800%
20) 2-Butanone-d5	4.642	46	122195	49.806	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	99.620%
24) Chloroform-d	5.067	84	69501	4.780	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.600%
26) 1,2-Dichloroethane-d4	5.703	65	39353	4.971	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.400%
32) Benzene-d6	5.732	84	131170	4.105	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.200%
36) 1,2-Dichloropropane-d6	6.694	67	44565	4.645	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.000%
41) Toluene-d8	7.899	98	100329	3.431	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	68.600%#
43) trans-1,3-Dichloroprop...	8.182	79	15120	3.589	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	71.800%
46) 2-Hexanone-d5	8.636	63	96900	38.843	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	77.680%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	41578	4.317	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.400%
66) 1,2-Dichlorobenzene-d4	12.195	152	37704	4.329	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	86.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU012822\
Data File : VU046894.D
Acq On : 28 Jan 2022 15:45
Operator : SY/MD
Sample : VIBLK026
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK026

Quant Time: Jan 29 00:41:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jan 29 00:38:36 2022
Response via : Initial Calibration

