

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU013122\
 Data File : VU046921.D
 Acq On : 31 Jan 2022 16:39
 Operator : SY/MD
 Sample : N1328-15 40X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0HN4

Quant Time: Feb 01 00:27:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Feb 01 00:23:47 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	98461	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	111844	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	45104	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	50857	6.096	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	122.000%	
7) Chloroethane-d5	1.919	69	37461	5.699	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	114.000%	
11) 1,1-Dichloroethene-d2	2.575	65	18496	5.573	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	111.400%	
20) 2-Butanone-d5	4.652	46	105148	42.541	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	85.080%	
24) Chloroform-d	5.067	84	72983	4.983	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	99.600%	
26) 1,2-Dichloroethane-d4	5.707	65	40987	5.139	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	102.800%	
32) Benzene-d6	5.732	84	146405	4.454	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	89.000%	
36) 1,2-Dichloropropane-d6	6.694	67	46060	4.667	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	93.400%	
41) Toluene-d8	7.903	98	112119	3.726	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	74.600%	
43) trans-1,3-Dichloroprop...	8.182	79	17084	3.942	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	78.800%	
46) 2-Hexanone-d5	8.639	63	97579	38.021	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	76.040%	
56) 1,1,2,2-Tetrachloroeth...	10.758	84	41614	4.200	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	84.000%	
66) 1,2-Dichlorobenzene-d4	12.195	152	38992	4.610	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	92.200%	
Target Compounds						
12) 1,1-Dichloroethene	2.588	96	7992	1.454	ug/L #	9
19) 1,1-Dichloroethane	3.867	63	1178	0.118	ug/L #	74
22) cis-1,2-Dichloroethene	4.671	96	28818	4.588	ug/L	98
29) 1,1,1-Trichloroethane	5.318	97	6330	0.628	ug/L	98
34) Trichloroethene	6.546	95	33098	4.841	ug/L	97
47) Tetrachloroethene	8.555	164	105464	18.967	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU013122\
Data File : VU046921.D
Acq On : 31 Jan 2022 16:39
Operator : SY/MD
Sample : N1328-15 40X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0HN4

Quant Time: Feb 01 00:27:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Feb 01 00:23:47 2022
Response via : Initial Calibration

