

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU013122\
 Data File : VU046927.D
 Acq On : 31 Jan 2022 19:06
 Operator : SY/MD
 Sample : N1273-15 10X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 COHL3

Quant Time: Feb 01 00:28:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Feb 01 00:23:47 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	112581	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	126003	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.816	152	54676	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	52086	5.460	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	109.200%
7) Chloroethane-d5	1.919	69	38350	5.102	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	102.000%
11) 1,1-Dichloroethene-d2	2.562	65	19709	5.194	ug/L	-0.01
Spiked Amount	5.000	Range	60 - 125	Recovery	=	103.800%
20) 2-Butanone-d5	4.655	46	123878	43.833	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	87.660%
24) Chloroform-d	5.067	84	74463	4.446	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.000%
26) 1,2-Dichloroethane-d4	5.703	65	42765	4.689	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.800%
32) Benzene-d6	5.729	84	155669	4.204	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.000%
36) 1,2-Dichloropropane-d6	6.694	67	48079	4.324	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	86.400%
41) Toluene-d8	7.899	98	123174	3.634	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	72.600%
43) trans-1,3-Dichloroprop...	8.182	79	18157	3.718	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	74.400%
46) 2-Hexanone-d5	8.658	63	111403	38.530	ug/L	0.02
Spiked Amount	50.000	Range	45 - 130	Recovery	=	77.060%
56) 1,1,2,2-Tetrachloroeth...	10.764	84	45972	4.118	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	82.400%
66) 1,2-Dichlorobenzene-d4	12.198	152	45719	4.459	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	89.200%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.575	96	9577	1.523	ug/L	# 1
19) 1,1-Dichloroethane	3.874	63	9065	0.793	ug/L	98
22) cis-1,2-Dichloroethene	4.671	96	37760	5.257	ug/L	99
34) Trichloroethene	6.543	95	52773	6.852	ug/L	98
47) Tetrachloroethene	8.555	164	175372	27.995	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU013122\
Data File : VU046927.D
Acq On : 31 Jan 2022 19:06
Operator : SY/MD
Sample : N1273-15 10X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0HL3

Quant Time: Feb 01 00:28:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Feb 01 00:23:47 2022
Response via : Initial Calibration

