

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU013122\
 Data File : VU046939.D
 Acq On : 01 Feb 2022 00:23
 Operator : SY/MD
 Sample : N1328-12DL 400X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0HN1DL

Quant Time: Feb 01 03:02:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Feb 01 01:39:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	103302	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	118139	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	48421	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	54640	6.242	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	124.800%	
7) Chloroethane-d5	1.919	69	42140	6.110	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	122.200%	
11) 1,1-Dichloroethene-d2	2.575	65	21111	6.063	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	121.200%	
20) 2-Butanone-d5	4.649	46	118353	45.640	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	91.280%	
24) Chloroform-d	5.070	84	81258	5.288	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	105.800%	
26) 1,2-Dichloroethane-d4	5.707	65	45928	5.488	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	109.800%	
32) Benzene-d6	5.732	84	163240	4.701	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	94.000%	
36) 1,2-Dichloropropane-d6	6.694	67	52092	4.997	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	100.000%	
41) Toluene-d8	7.899	98	125869	3.961	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	79.200%	
43) trans-1,3-Dichloroprop...	8.182	79	18753	4.096	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	82.000%	
46) 2-Hexanone-d5	8.639	63	111754	41.224	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	82.440%	
56) 1,1,2,2-Tetrachloroeth...	10.758	84	49088	4.690	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	93.800%	
66) 1,2-Dichlorobenzene-d4	12.195	152	43185	4.756	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	95.200%	
Target Compounds						Qvalue
19) 1,1-Dichloroethane	3.877	63	1336	0.127	ug/L	# 89
22) cis-1,2-Dichloroethene	4.671	96	10574	1.604	ug/L	90
47) Tetrachloroethene	8.555	164	28896	4.920	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU013122\
Data File : VU046939.D
Acq On : 01 Feb 2022 00:23
Operator : SY/MD
Sample : N1328-12DL 400X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0HN1DL

Quant Time: Feb 01 03:02:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Feb 01 01:39:38 2022
Response via : Initial Calibration

