

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU013122\
 Data File : VU046944.D
 Acq On : 01 Feb 2022 02:25
 Operator : SY/MD
 Sample : N1328-18DL 1000X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0HN7DL

Quant Time: Feb 01 03:03:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Feb 01 01:39:38 2022
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene		6.253	114	103717	5.000	ug/L	0.00
28) Chlorobenzene-d5		9.420	117	118705	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4		11.812	152	47422	5.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3		1.604	65	53045	6.036	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	120.800%	
7) Chloroethane-d5		1.922	69	40694	5.877	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	117.600%	
11) 1,1-Dichloroethene-d2		2.575	65	20552	5.879	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	117.600%	
20) 2-Butanone-d5		4.652	46	111923	42.988	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	85.980%	
24) Chloroform-d		5.070	84	77617	5.031	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.600%	
26) 1,2-Dichloroethane-d4		5.706	65	43630	5.193	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.800%	
32) Benzene-d6		5.732	84	151540	4.344	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.800%	
36) 1,2-Dichloropropane-d6		6.693	67	48831	4.662	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.200%	
41) Toluene-d8		7.899	98	118133	3.699	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	74.000%	
43) trans-1,3-Dichloroprop...		8.182	79	17551	3.815	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	76.400%	
46) 2-Hexanone-d5		8.639	63	109407	40.166	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	80.340%	
56) 1,1,2,2-Tetrachloroeth...		10.758	84	46228	4.396	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	88.000%	
66) 1,2-Dichlorobenzene-d4		12.195	152	41157	4.628	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	92.600%	
Target Compounds							
12) 1,1-Dichloroethene		2.587	96	6844	1.182	ug/L	# 1
22) cis-1,2-Dichloroethene		4.674	96	672	0.102	ug/L	95
34) Trichloroethene		6.549	95	15208	2.096	ug/L	94
47) Tetrachloroethene		8.555	164	31545	5.345	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU013122\
Data File : VU046944.D
Acq On : 01 Feb 2022 02:25
Operator : SY/MD
Sample : N1328-18DL 1000X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0HN7DL

Quant Time: Feb 01 03:03:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Feb 01 01:39:38 2022
Response via : Initial Calibration

