

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU013122\
 Data File : VU046946.D
 Acq On : 01 Feb 2022 03:14
 Operator : SY/MD
 Sample : N1273-16DL 500X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0HL4DL

Quant Time: Feb 01 03:46:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Feb 01 01:39:38 2022
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene		6.253	114	105141	5.000	ug/L	0.00
28) Chlorobenzene-d5		9.420	117	118682	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4		11.812	152	48188	5.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3		1.604	65	54261	6.090	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	121.800%	
7) Chloroethane-d5		1.919	69	42283	6.024	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	120.400%	
11) 1,1-Dichloroethene-d2		2.575	65	21335	6.020	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	120.400%	
20) 2-Butanone-d5		4.652	46	115635	43.812	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	87.620%	
24) Chloroform-d		5.070	84	84619	5.410	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.200%	
26) 1,2-Dichloroethane-d4		5.706	65	46530	5.463	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	109.200%	
32) Benzene-d6		5.732	84	162964	4.672	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.400%	
36) 1,2-Dichloropropane-d6		6.694	67	52534	5.016	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.400%	
41) Toluene-d8		7.899	98	128942	4.039	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.800%	
43) trans-1,3-Dichloroprop...		8.182	79	19541	4.249	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.000%	
46) 2-Hexanone-d5		8.639	63	117709	43.222	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	86.440%	
56) 1,1,2,2-Tetrachloroeth...		10.758	84	49976	4.753	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.000%	
66) 1,2-Dichlorobenzene-d4		12.195	152	42658	4.720	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	94.400%	
Target Compounds							
12) 1,1-Dichloroethene		2.588	96	12765	2.174	ug/L	Qvalue # 37
22) cis-1,2-Dichloroethene		4.671	96	2159	0.322	ug/L	91
29) 1,1,1-Trichloroethane		5.321	97	6265	0.586	ug/L	97
34) Trichloroethene		6.546	95	57266	7.893	ug/L	98
47) Tetrachloroethene		8.555	164	25554	4.331	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU013122\
Data File : VU046946.D
Acq On : 01 Feb 2022 03:14
Operator : SY/MD
Sample : N1273-16DL 500X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0HL4DL

Quant Time: Feb 01 03:46:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Feb 01 01:39:38 2022
Response via : Initial Calibration

