

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU013122\
 Data File : VU046947.D
 Acq On : 01 Feb 2022 03:38
 Operator : SY/MD
 Sample : N1273-17DL 500X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0HL5DL

Quant Time: Feb 01 03:53:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Feb 01 01:39:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	101302	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	115584	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	45363	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	51740	6.028	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	120.600%
7) Chloroethane-d5	1.922	69	39769	5.880	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	117.600%
11) 1,1-Dichloroethene-d2	2.575	65	19940	5.840	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	116.800%
20) 2-Butanone-d5	4.652	46	86290	33.933	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	67.860%
24) Chloroform-d	5.070	84	77876	5.168	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.400%
26) 1,2-Dichloroethane-d4	5.706	65	44090	5.373	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.400%
32) Benzene-d6	5.732	84	154165	4.538	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.800%
36) 1,2-Dichloropropane-d6	6.694	67	49685	4.871	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.400%
41) Toluene-d8	7.903	98	121308	3.901	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.000%
43) trans-1,3-Dichloroprop...	8.182	79	18027	4.025	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.400%
46) 2-Hexanone-d5	8.639	63	112325	42.351	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	84.700%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	46368	4.528	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.600%
66) 1,2-Dichlorobenzene-d4	12.195	152	40851	4.802	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.000%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.588	96	11912	2.106	ug/L	# 22
22) cis-1,2-Dichloroethene	4.674	96	2221	0.344	ug/L	94
29) 1,1,1-Trichloroethane	5.317	97	5871	0.564	ug/L	97
34) Trichloroethene	6.546	95	50168	7.100	ug/L	98
47) Tetrachloroethene	8.555	164	23253	4.047	ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU013122\
Data File : VU046947.D
Acq On : 01 Feb 2022 03:38
Operator : SY/MD
Sample : N1273-17DL 500X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0HL5DL

Quant Time: Feb 01 03:53:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Feb 01 01:39:38 2022
Response via : Initial Calibration

