

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU013122\
 Data File : VU046932.D
 Acq On : 31 Jan 2022 21:33
 Operator : SY/MD
 Sample : VU0131WBL02
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK058

Quant Time: Feb 01 01:39:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Feb 01 01:39:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	100732	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	113719	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	46537	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	50026	5.861	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	117.200%
7) Chloroethane-d5	1.919	69	37859	5.629	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	112.600%
11) 1,1-Dichloroethene-d2	2.575	65	19118	5.631	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	112.600%
20) 2-Butanone-d5	4.649	46	112026	44.302	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	88.600%
24) Chloroform-d	5.070	84	74460	4.969	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.400%
26) 1,2-Dichloroethane-d4	5.707	65	42732	5.237	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.800%
32) Benzene-d6	5.732	84	147937	4.426	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.600%
36) 1,2-Dichloropropane-d6	6.694	67	47389	4.722	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.400%
41) Toluene-d8	7.903	98	114664	3.748	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	75.000%
43) trans-1,3-Dichloroprop...	8.182	79	17107	3.882	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	77.600%
46) 2-Hexanone-d5	8.639	63	100692	38.587	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	77.180%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	43680	4.336	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.800%
66) 1,2-Dichlorobenzene-d4	12.195	152	39484	4.524	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	90.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU013122\
Data File : VU046932.D
Acq On : 31 Jan 2022 21:33
Operator : SY/MD
Sample : VU0131WBL02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK058

Quant Time: Feb 01 01:39:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Feb 01 01:39:38 2022
Response via : Initial Calibration

