

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020222\
 Data File : VU047004.D
 Acq On : 02 Feb 2022 16:15
 Operator : SY/MD
 Sample : N1349-11
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0HQ0

Quant Time: Feb 03 03:01:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Feb 03 02:56:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	95828	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	108813	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	45909	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	43047	5.301	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	106.000%
7) Chloroethane-d5	1.919	69	35536	5.554	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	111.000%
11) 1,1-Dichloroethene-d2	2.575	65	15816	4.897	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	98.000%
20) 2-Butanone-d5	4.629	46	101158	42.051	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	84.100%
24) Chloroform-d	5.066	84	72618	5.094	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.800%
26) 1,2-Dichloroethane-d4	5.706	65	38362	4.942	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.800%
32) Benzene-d6	5.732	84	140760	4.401	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.000%
36) 1,2-Dichloropropane-d6	6.693	67	45450	4.733	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.600%
41) Toluene-d8	7.902	98	107176	3.661	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	73.200%
43) trans-1,3-Dichloroprop...	8.182	79	15748	3.735	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	74.600%
46) 2-Hexanone-d5	8.636	63	79674	31.909	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	63.820%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	37163	3.855	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	77.000%
66) 1,2-Dichlorobenzene-d4	12.195	152	41970	4.875	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.400%
Target Compounds						
					Qvalue	
3) Chloromethane	1.526	50	3739	0.504	ug/L	90
12) 1,1-Dichloroethene	2.587	96	14315	2.675	ug/L	77
13) Acetone	2.639	43	7427	5.677	ug/L	97
29) 1,1,1-Trichloroethane	5.321	97	5044	0.515	ug/L #	89
33) Benzene	5.780	78	11204	0.437	ug/L	100
34) Trichloroethene	6.546	95	126414	19.005	ug/L	99
42) Toluene	7.973	91	17159	0.560	ug/L	98
47) Tetrachloroethene	8.555	164	28036	5.182	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU020222\
Data File : VU047004.D
Acq On : 02 Feb 2022 16:15
Operator : SY/MD
Sample : N1349-11
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0HQ0

Quant Time: Feb 03 03:01:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Feb 03 02:56:36 2022
Response via : Initial Calibration

