

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021122\
 Data File : VU047105.D
 Acq On : 11 Feb 2022 10:51
 Operator : SY/MD
 Sample : VU0211WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK066

Quant Time: Feb 11 11:51:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM021022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Feb 11 01:13:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	475754	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	441840	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	227808	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	196134	45.393	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	90.780%
7) Chloroethane-d5	1.922	69	151263	47.225	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	94.460%
11) 1,1-Dichloroethene-d2	2.575	63	259251	36.207	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.420%
21) 2-Butanone-d5	4.639	46	417047	124.453	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	124.450%
24) Chloroform-d	5.070	84	321487	45.832	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.660%
26) 1,2-Dichloroethane-d4	5.706	65	217891	48.396	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.800%
32) Benzene-d6	5.732	84	699957	47.651	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.300%
36) 1,2-Dichloropropane-d6	6.694	67	228776	48.025	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.060%
41) Toluene-d8	7.899	98	659591	47.161	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.320%
43) trans-1,3-Dichloroprop...	8.179	79	117784	47.182	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	94.360%
47) 2-Hexanone-d5	8.636	63	304174	111.457	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	111.460%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	334129	48.104	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	96.200%
66) 1,2-Dichlorobenzene-d4	12.195	152	253177	46.860	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.720%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021122\
Data File : VU047105.D
Acq On : 11 Feb 2022 10:51
Operator : SY/MD
Sample : VU0211WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK066

Quant Time: Feb 11 11:51:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM021022WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Feb 11 01:13:45 2022
Response via : Initial Calibration

