

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061622\
 Data File : VU049213.D
 Acq On : 16 Jun 2022 18:44
 Operator : SY/MD
 Sample : VU0616WBL02
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK090

Quant Time: Jun 17 02:32:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jun 17 02:27:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	121340	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	113700	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	55659	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	40250	3.796	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	76.000%	
7) Chloroethane-d5	1.912	69	32327	4.504	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	90.000%	
11) 1,1-Dichloroethene-d2	2.565	65	17068	4.002	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	80.000%	
20) 2-Butanone-d5	4.636	46	170136	60.099	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	120.200%	
24) Chloroform-d	5.067	84	81044	4.872	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	97.400%	
26) 1,2-Dichloroethane-d4	5.703	65	47254	5.178	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	103.600%	
32) Benzene-d6	5.729	84	160522	4.686	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	93.800%	
36) 1,2-Dichloropropane-d6	6.694	67	55658	4.788	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	95.800%	
41) Toluene-d8	7.899	98	139215	4.514	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	90.200%	
43) trans-1,3-Dichloroprop...	8.182	79	21683	4.198	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	84.000%	
46) 2-Hexanone-d5	8.636	63	114267	54.846	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	109.700%	
56) 1,1,2,2-Tetrachloroeth...	10.758	84	47164	5.067	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	101.400%	
66) 1,2-Dichlorobenzene-d4	12.195	152	52330	4.900	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	98.000%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061622\
Data File : VU049213.D
Acq On : 16 Jun 2022 18:44
Operator : SY/MD
Sample : VU0616WBL02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK090

Quant Time: Jun 17 02:32:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jun 17 02:27:54 2022
Response via : Initial Calibration

