

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061622\
 Data File : VU049218.D
 Acq On : 16 Jun 2022 20:46
 Operator : SY/MD
 Sample : N3279-02
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0BG3

Quant Time: Jun 17 02:33:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jun 17 02:27:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	125006	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	115271	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	57112	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	35144	3.217	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	64.400%	
7) Chloroethane-d5	1.912	69	28753	3.888	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	77.800%	
11) 1,1-Dichloroethene-d2	2.568	65	14747	3.357	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	67.200%	
20) 2-Butanone-d5	4.639	46	167093	57.293	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	114.580%	
24) Chloroform-d	5.067	84	72993	4.259	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	85.200%	
26) 1,2-Dichloroethane-d4	5.703	65	42494	4.520	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	90.400%	
32) Benzene-d6	5.729	84	141187	4.066	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	81.400%	
36) 1,2-Dichloropropane-d6	6.690	67	51225	4.346	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	87.000%	
41) Toluene-d8	7.899	98	119599	3.825	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	76.400%	
43) trans-1,3-Dichloroprop...	8.182	79	19660	3.755	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	75.000%	
46) 2-Hexanone-d5	8.636	63	115066	54.477	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	108.960%	
56) 1,1,2,2-Tetrachloroeth...	10.758	84	44996	4.768	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	95.400%	
66) 1,2-Dichlorobenzene-d4	12.195	152	48255	4.403	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	88.000%	
Target Compounds						
						Qvalue
9) Trichlorofluoromethane	2.137	101	726	0.060	ug/L	88
14) Carbon disulfide	2.793	76	18887	0.813	ug/L	100
34) Trichloroethene	6.546	95	4522	0.516	ug/L	97
42) Toluene	7.970	91	25659	0.705	ug/L	98
52) Ethylbenzene	9.574	91	3573	0.090	ug/L	84
53) m,p-Xylene	9.693	106	6975	0.472	ug/L	91
63) 1,2,4-Trimethylbenzene	11.471	105	4398	0.129	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061622\
Data File : VU049218.D
Acq On : 16 Jun 2022 20:46
Operator : SY/MD
Sample : N3279-02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0BG3

Quant Time: Jun 17 02:33:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jun 17 02:27:54 2022
Response via : Initial Calibration

