

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061622\
 Data File : VU049230.D
 Acq On : 17 Jun 2022 01:31
 Operator : SY/MD
 Sample : N3320-05
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BGJ79

Quant Time: Jun 17 02:36:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jun 17 02:27:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	118431	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	107150	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	47704	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	39694	3.836	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	76.800%
7) Chloroethane-d5	1.912	69	31987	4.566	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	91.400%
11) 1,1-Dichloroethene-d2	2.565	65	16598	3.988	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	79.800%
20) 2-Butanone-d5	4.636	46	162822	58.928	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	117.860%
24) Chloroform-d	5.063	84	78107	4.811	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.200%
26) 1,2-Dichloroethane-d4	5.703	65	45885	5.152	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.000%
32) Benzene-d6	5.729	84	155652	4.822	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.400%
36) 1,2-Dichloropropane-d6	6.690	67	54910	5.012	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.200%
41) Toluene-d8	7.899	98	134476	4.627	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.600%
43) trans-1,3-Dichloroprop...	8.182	79	17111	3.516	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	70.400%
46) 2-Hexanone-d5	8.636	63	110460	56.260	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	112.520%
56) 1,1,2,2-Tetrachloroeth...	10.755	84	45753	5.216	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	104.400%
66) 1,2-Dichlorobenzene-d4	12.192	152	45438	4.964	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.200%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.581	96	49420	7.421	ug/L	85
13) Acetone	2.652	43	10421	7.620	ug/L	88
14) Carbon disulfide	2.797	76	3061	0.139	ug/L	95
19) 1,1-Dichloroethane	3.871	63	232793	16.002	ug/L	97
21) 2-Butanone	4.726	43	8386	3.216	ug/L	76
22) cis-1,2-Dichloroethene	4.671	96	1077	0.131	ug/L	83
27) 1,2-Dichloroethane	5.797	62	5986	0.622	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061622\
Data File : VU049230.D
Acq On : 17 Jun 2022 01:31
Operator : SY/MD
Sample : N3320-05
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BGJ79

Quant Time: Jun 17 02:36:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jun 17 02:27:54 2022
Response via : Initial Calibration

