

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061822\
 Data File : VU049292.D
 Acq On : 18 Jun 2022 19:32
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK074

Quant Time: Jun 20 00:15:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM061022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Jun 20 00:11:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	261346	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	251490	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	114632	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	76530	47.534	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	95.060%
7) Chloroethane-d5	1.909	69	62175	52.595	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	105.200%
11) 1,1-Dichloroethene-d2	2.565	63	105731	34.803	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	69.600%
21) 2-Butanone-d5	4.620	46	189676	107.453	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	107.450%
24) Chloroform-d	5.063	84	162551	47.328	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.660%
26) 1,2-Dichloroethane-d4	5.700	65	110305	49.572	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.140%
32) Benzene-d6	5.726	84	317346	50.888	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.780%
36) 1,2-Dichloropropane-d6	6.690	67	115164	54.934	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	109.860%
41) Toluene-d8	7.896	98	280314	50.969	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.940%
43) trans-1,3-Dichloroprop...	8.179	79	48430	49.104	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	98.200%
47) 2-Hexanone-d5	8.632	63	119178	108.619	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	108.620%
56) 1,1,2,2-Tetrachloroeth...	10.755	84	168630	48.160	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	96.320%
66) 1,2-Dichlorobenzene-d4	12.192	152	107486	53.890	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	107.780%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU061822\
Data File : VU049292.D
Acq On : 18 Jun 2022 19:32
Operator : SY/MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK074

Quant Time: Jun 20 00:15:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM061022WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Jun 20 00:11:46 2022
Response via : Initial Calibration

