

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092322\
 Data File : VU050927.D
 Acq On : 23 Sep 2022 18:18
 Operator : JC/MD
 Sample : N4651-15
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Sep 24 00:25:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM092022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Sep 24 00:22:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	397658	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	395704	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	139541	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	164639	39.961	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	79.920%	
7) Chloroethane-d5	1.909	69	146653	49.980	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	99.960%	
11) 1,1-Dichloroethene-d2	2.565	63	211996	33.069	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	66.140%	
21) 2-Butanone-d5	4.626	46	322402	121.210	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	121.210%	
24) Chloroform-d	5.060	84	315540	48.871	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	97.740%	
26) 1,2-Dichloroethane-d4	5.700	65	210918	53.007	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	106.020%	
32) Benzene-d6	5.726	84	625199	47.968	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	95.940%	
36) 1,2-Dichloropropane-d6	6.690	67	220547	50.131	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	100.260%	
41) Toluene-d8	7.896	98	510370	44.797	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	89.600%	
43) trans-1,3-Dichloroprop...	8.179	79	89356	46.688	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	93.380%	
47) 2-Hexanone-d5	8.632	63	176223	97.199	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	97.200%	
56) 1,1,2,2-Tetrachloroeth...	10.754	84	328296	48.945	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	97.900%	
66) 1,2-Dichlorobenzene-d4	12.192	152	168978	57.837	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	115.680%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092322\
Data File : VU050927.D
Acq On : 23 Sep 2022 18:18
Operator : JC/MD
Sample : N4651-15
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Sep 24 00:25:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM092022WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Sep 24 00:22:43 2022
Response via : Initial Calibration

