

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092622\
 Data File : VU050940.D
 Acq On : 26 Sep 2022 12:29
 Operator : JC/MD
 Sample : N4747-07
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 E45A4

Quant Time: Sep 26 20:54:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM092022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Sep 26 20:52:47 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	367629	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	371074	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	150807	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	150902	39.618	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	79.240%
7) Chloroethane-d5	1.912	69	129728	47.823	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	95.640%
11) 1,1-Dichloroethene-d2	2.568	63	199248	33.619	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	67.240%
21) 2-Butanone-d5	4.620	46	286748	116.611	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	116.610%
24) Chloroform-d	5.063	84	295442	49.496	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.000%
26) 1,2-Dichloroethane-d4	5.700	65	193518	52.607	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.220%
32) Benzene-d6	5.726	84	575728	47.104	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.200%
36) 1,2-Dichloropropane-d6	6.690	67	204544	49.579	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.160%
41) Toluene-d8	7.896	98	483258	45.232	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.460%
43) trans-1,3-Dichloroprop...	8.179	79	83695	46.633	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	93.260%
47) 2-Hexanone-d5	8.632	63	156929	92.302	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	92.300%
56) 1,1,2,2-Tetrachloroeth...	10.754	84	310392	49.347	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	98.700%
66) 1,2-Dichlorobenzene-d4	12.192	152	175821	55.683	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	111.360%
Target Compounds						
					Qvalue	
5) Vinyl chloride	1.604	62	6670	1.540	ug/L	# 61
17) trans-1,2-Dichloroethene	3.362	96	1887	0.651	ug/L	82
20) cis-1,2-Dichloroethene	4.668	96	52975	16.493	ug/L	95
34) Trichloroethene	6.542	95	254501	80.231	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092622\
Data File : VU050940.D
Acq On : 26 Sep 2022 12:29
Operator : JC/MD
Sample : N4747-07
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E45A4

Quant Time: Sep 26 20:54:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM092022WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Sep 26 20:52:47 2022
Response via : Initial Calibration

