

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092722\
 Data File : VU050953.D
 Acq On : 27 Sep 2022 13:47
 Operator : JC/MD
 Sample : N4854-01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F5J01

Quant Time: Sep 28 02:01:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM092022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Sep 28 01:59:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	332248	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	353719	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	145523	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	111617	32.425	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	64.860%	
7) Chloroethane-d5	1.913	69	105023	42.839	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	85.680%	
11) 1,1-Dichloroethene-d2	2.565	63	151658	28.314	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	56.620%#	
21) 2-Butanone-d5	4.620	46	236859	106.580	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	106.580%	
24) Chloroform-d	5.064	84	260307	48.254	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	96.500%	
26) 1,2-Dichloroethane-d4	5.700	65	172655	51.934	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	103.860%	
32) Benzene-d6	5.729	84	485723	41.690	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	83.380%	
36) 1,2-Dichloropropane-d6	6.690	67	178002	45.263	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	90.520%	
41) Toluene-d8	7.899	98	401965	39.470	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	78.940%#	
43) trans-1,3-Dichloroprop...	8.179	79	71339	41.699	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	83.400%	
47) 2-Hexanone-d5	8.633	63	145167	89.573	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	89.570%	
56) 1,1,2,2-Tetrachloroeth...	10.755	84	276157	46.059	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	92.120%	
66) 1,2-Dichlorobenzene-d4	12.192	152	156406	51.333	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	102.660%	
Target Compounds						
					Qvalue	
18) Methyl tert-butyl Ether	3.366	73	5846	0.732	ug/L #	77
34) Trichloroethene	6.543	95	116030	38.373	ug/L	95
46) Tetrachloroethene	8.555	164	1392359	640.774	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092722\
Data File : VU050953.D
Acq On : 27 Sep 2022 13:47
Operator : JC/MD
Sample : N4854-01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F5J01

Quant Time: Sep 28 02:01:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM092022WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Sep 28 01:59:42 2022
Response via : Initial Calibration

