

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092722\
 Data File : VU050955.D
 Acq On : 27 Sep 2022 14:34
 Operator : JC/MD
 Sample : N4854-13
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F5J55

Quant Time: Sep 28 02:02:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM092022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Sep 28 01:59:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	328796	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	349045	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	145627	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	113994	33.463	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	66.920%	
7) Chloroethane-d5	1.912	69	107249	44.206	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	88.420%	
11) 1,1-Dichloroethene-d2	2.565	63	155099	29.260	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	58.520%#	
21) 2-Butanone-d5	4.620	46	264905	120.452	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	120.450%	
24) Chloroform-d	5.063	84	269436	50.470	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	100.940%	
26) 1,2-Dichloroethane-d4	5.700	65	178303	54.196	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	108.400%	
32) Benzene-d6	5.726	84	505613	43.978	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	87.960%	
36) 1,2-Dichloropropane-d6	6.690	67	180402	46.487	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	92.980%	
41) Toluene-d8	7.896	98	416367	41.431	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	82.860%	
43) trans-1,3-Dichloroprop...	8.176	79	74076	43.878	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	87.760%	
47) 2-Hexanone-d5	8.632	63	141175	88.277	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	88.280%	
56) 1,1,2,2-Tetrachloroeth...	10.754	84	284531	48.091	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	96.180%	
66) 1,2-Dichlorobenzene-d4	12.192	152	167721	55.007	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	110.020%	
Target Compounds						Qvalue
13) Acetone	2.629	43	7924	4.255	ug/L	97
46) Tetrachloroethene	8.552	164	29846	13.919	ug/L	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092722\
Data File : VU050955.D
Acq On : 27 Sep 2022 14:34
Operator : JC/MD
Sample : N4854-13
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F5J55

Quant Time: Sep 28 02:02:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM092022WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Sep 28 01:59:42 2022
Response via : Initial Calibration





#13

Acetone

Concen: 4.255 ug/L

RT: 2.629 min Scan# 401

Delta R.T. 0.006 min

Lab File: VU050955.D

Acq: 27 Sep 2022 14:34

Instrument :

MSVOA_U

ClientSampleId :





F5J55

Tgt Ion: 43 Resp: 7924

Ion Ratio Lower Upper

43 100

58 34.2 0.0 71.4



#46

Tetrachloroethene

Concen: 13.919 ug/L

RT: 8.552 min Scan# 2243

Delta R.T. -0.000 min

Lab File: VU050955.D

Acq: 27 Sep 2022 14:34

Tgt Ion:164 Resp: 29846

Ion Ratio Lower Upper

164 100

129 96.5 68.5 127.3

131 93.7 66.1 122.7

166 125.9 88.8 164.8

