

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU093022\
 Data File : VU051053.D
 Acq On : 30 Sep 2022 23:09
 Operator : JC/MD
 Sample : VU0930WBL02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK194

Quant Time: Oct 01 03:05:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM092922WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Oct 01 03:04:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	479921	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	462583	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	174360	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	197499	40.606	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	81.220%	
7) Chloroethane-d5	1.912	69	177242	39.230	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	78.460%	
11) 1,1-Dichloroethene-d2	2.565	63	250639	34.886	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	69.780%	
21) 2-Butanone-d5	4.632	46	348629	110.516	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	110.520%	
24) Chloroform-d	5.060	84	357352	45.603	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	91.200%	
26) 1,2-Dichloroethane-d4	5.700	65	236604	48.782	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	97.560%	
32) Benzene-d6	5.726	84	739271	54.574	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	109.140%	
36) 1,2-Dichloropropane-d6	6.687	67	246132	55.995	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	111.980%	
41) Toluene-d8	7.896	98	621712	49.675	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	99.340%	
43) trans-1,3-Dichloroprop...	8.179	79	101816	48.563	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	97.120%	
47) 2-Hexanone-d5	8.636	63	187344	105.553	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	105.550%	
56) 1,1,2,2-Tetrachloroeth...	10.754	84	360439	46.919	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	93.840%	
66) 1,2-Dichlorobenzene-d4	12.192	152	196867	52.853	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	105.700%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU093022\
Data File : VU051053.D
Acq On : 30 Sep 2022 23:09
Operator : JC/MD
Sample : VU0930WBL02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK194

Quant Time: Oct 01 03:05:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM092922WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Oct 01 03:04:59 2022
Response via : Initial Calibration

