

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100722\
 Data File : VU051222.D
 Acq On : 07 Oct 2022 16:54
 Operator : JC/MD
 Sample : N4873-02
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EXZ50

Quant Time: Oct 08 04:26:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR100722WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Oct 08 00:50:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	208172	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	227900	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	82923	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	68833	4.099	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	82.000%
7) Chloroethane-d5	1.909	69	62903	4.360	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	87.200%
11) 1,1-Dichloroethene-d2	2.562	65	25806	3.888	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	77.800%
20) 2-Butanone-d5	4.620	46	177713	48.069	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	96.140%
24) Chloroform-d	5.060	84	135809	4.338	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.800%
26) 1,2-Dichloroethane-d4	5.700	65	73099	4.528	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.600%
32) Benzene-d6	5.726	84	259187	4.134	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.600%
36) 1,2-Dichloropropane-d6	6.690	67	92040	4.323	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	86.400%
41) Toluene-d8	7.896	98	206648	3.684	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	73.600%
43) trans-1,3-Dichloroprop...	8.179	79	28483	4.072	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.400%
46) 2-Hexanone-d5	8.632	63	94736	40.182	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	80.360%
56) 1,1,2,2-Tetrachloroeth...	10.754	84	75235	4.423	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	88.400%
66) 1,2-Dichlorobenzene-d4	12.192	152	66190	4.592	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	91.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100722\
Data File : VU051222.D
Acq On : 07 Oct 2022 16:54
Operator : JC/MD
Sample : N4873-02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EXZ50

Quant Time: Oct 08 04:26:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR100722WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Oct 08 00:50:53 2022
Response via : Initial Calibration

